



Updated Open-air Lagoon Show with Pyrotechnic Effects

Noise Monitoring Report (September
2024)

PREPARED FOR



Ocean Park Corporation

DATE

4 October 2024

REFERENCE

0540005



DOCUMENT DETAILS

The details entered below are automatically shown on the cover and the main page footer. PLEASE NOTE: This table must NOT be removed from this document.

| | |
|-------------------|---|
| DOCUMENT TITLE | Updated Open-air Lagoon Show with Pyrotechnic Effects |
| DOCUMENT SUBTITLE | Noise Monitoring Report (September 2024) |
| PROJECT NUMBER | 0540005 |
| Date | 4 October 2024 |
| Version | 01 |
| Author | Various |
| Client name | Ocean Park Corporation |

DOCUMENT HISTORY

| | | | | ERM APPROVAL TO ISSUE | | |
|---------|----------|---------|-------------|-----------------------|------------|----------|
| VERSION | REVISION | AUTHOR | REVIEWED BY | NAME | DATE | COMMENTS |
| Version | 1.0 | Various | MT | TF | 04.10.2024 | - |
| | | | | | | |
| | | | | | | |

SIGNATURE PAGE

Updated Open-air Lagoon Show with Pyrotechnic Effects

Noise Monitoring Report (September 2024)

0540005



Terence Fong

Partner

ERM-Hong Kong, Limited
2509, 25/F One Harbourfront
18 Tak Fung Street
Hung Hom, Kowloon
Hong Kong
T +852 2271 3000

© Copyright 2023 by The ERM International Group Limited and/or its affiliates ('ERM'). All Rights Reserved.
No part of this work may be reproduced or transmitted in any form or by any means, without prior written permission of ERM.





Environmental Permit No. EP-249/2006/D
Ocean Park Master Redevelopment Project
Environmental Team Leader Certification

Reference Document/Plan

| | |
|--|--|
| Document/ Plan to be Certified/ Verified : | Noise Monitoring Report (September 2024) |
| Date of Report: | 4 October 2024 |

Reference EP Condition

| | |
|---|-----|
| Environmental Permit Condition: | 3.4 |
| Four hard copies and one electronic copy of the monthly EM&A Reports for the construction and operation stages shall be submitted to the Director within two weeks after the end of the reporting month. The monthly EM&A Reports shall include a summary of all non-compliance with the recommendations in the EIA Report or this Permit. The submissions shall be certified by the ET Leader and verified by the IEC as complied with the requirements as set out in the EM&A Manual before submission to the Director. Additional copies of the submission shall be provided upon request by the Director. | |

ETL Verification

I hereby verify that the above referenced document/~~plan~~ complies with the above referenced condition of EP-249/2006/D.

Ms Mandy To
Environmental Team Leader

Date: 4 October 2024

Our ref: 0540005_ETL Verification Cert_20241004.docx

Ocean Park Master Redevelopment Project

Environmental Permit No. EP-249/2006/D - Condition 3.4

**Updated Open-air Lagoon Show with Pyrotechnic Effects
Monthly Noise Monitoring Report
(September 2024)**

Submitted by ERM-Hong Kong, Limited dated 4 October 2024


This is to verify that

**Updated Open-air Lagoon Show with Pyrotechnic Effects
Monthly Noise Monitoring Report
(September 2024)**

Submitted by ERM-Hong Kong, Limited dated 4 October 2024

Has been verified by the undersigned.

Signed



Ir Eric Ching
Independent Environmental Checker (IEC)
Retained by Ocean Park Corporation
pursuant to Environmental Permit No. EP-249/2006/D

Date

7 October 2024

CONTENTS

| | | |
|-----------|-------------------------------|----------|
| 1. | BACKGROUND | 1 |
| 1.1 | PURPOSE OF THE REPORT | 1 |
| 1.2 | STRUCTURE OF THE REPORT | 1 |
| 2. | NOISE MONITORING | 2 |
| 2.1 | INTRODUCTION | 2 |
| 2.2 | NOISE MONITORING REQUIREMENTS | 2 |
| 2.3 | MONITORING LOCATIONS | 2 |
| 2.4 | MONITORING PARAMETERS | 2 |
| 2.5 | MONITORING FREQUENCY | 3 |
| 2.6 | MONITORING METHODOLOGY | 3 |
| 2.7 | COMPLIANCE ASSESSMENT | 4 |
| 2.8 | RESULTS OF NOISE MONITORING | 5 |
| 2.9 | SUMMARY OF NOISE EXCEEDANCES | 5 |
| 3. | CONCLUSION | 6 |

| | |
|------------|---|
| APPENDIX A | CALIBRATION CERTIFICATES OF THE NOISE MEASUREMENT EQUIPMENT |
| APPENDIX B | RESULTS OF NOISE MONITORING |
| APPENDIX C | PHOTOGRAPHS OF THE MONITORING STATIONS |

LIST OF TABLES

| | | |
|---------|---|---|
| TABLE 1 | NOISE MONITORING STATIONS | 2 |
| TABLE 2 | SCHEDULE OF LAGOON NIGHT SHOW IN SEPTEMBER 2024 | 3 |
| TABLE 3 | NOISE MONITORING SCHEDULE IN SEPTEMBER 2024 | 3 |
| TABLE 4 | NOISE MONITORING EQUIPMENT | 3 |
| TABLE 5 | ACTION AND LIMIT LEVELS FOR ENTERTAINMENT NOISE | 4 |
| TABLE 6 | COMPLIANCE OF NOISE MONITORING | 5 |

LIST OF FIGURES

| | |
|----------|----------------------------|
| FIGURE 1 | NOISE MONITORING LOCATIONS |
|----------|----------------------------|

1. BACKGROUND

ERM-Hong Kong, Limited (ERM) has been appointed by Ocean Park Corporation (OPC) to undertake noise monitoring of the Updated Open-air Lagoon Show with Pyrotechnic Effects under the "Repositioning and Long Term Operation Plan of Ocean Park" (the Project) with update of design in audio system and use of Pyrotechnic Special Effect Materials (PSEM), as presented in the *Noise Review Study Report 2024 (NRR 2024)* deposited to the Environmental Protection Department (EPD) on 6 June 2024.

1.1 PURPOSE OF THE REPORT

The Updated Open-air Lagoon Show with Pyrotechnic Effects, including *Soul of the Ocean (SOTO)* and *Vision of Hong Kong (VHK)*, commenced on 6 July 2024. As stated in *NRR 2024*, noise monitoring is proposed to be conducted on the first day of the updated lagoon night, as well as by the end of July, August and September 2024. The necessity and frequency of further noise monitoring should be further reviewed afterwards.

This is the noise monitoring report which summarises the impact monitoring results for the show held in **September 2024**.

1.2 STRUCTURE OF THE REPORT

After this introductory section, the remainder of this report is arranged as follows:

Section 2: describes the noise monitoring methodology, presents and discusses the monitoring results

Section 3: presents an overall conclusion of the noise monitoring

2. NOISE MONITORING

2.1 INTRODUCTION

Noise monitoring has been carried out following the requirements given in Condition 3.1 of the Environmental Permit (EP-249/2006/D) and the updated EM&A Manual. The requirements and results are detailed in the following sections.

2.2 NOISE MONITORING REQUIREMENTS

Operational phase noise monitoring during the first operational year of the lagoon night show was completed in accordance with the approved EIA report for “*Relocation and Long Term Operation Plan of Ocean Park*” and EM&A Manual. Nevertheless, due to update in design of audio system and the use of PSEM, as stated in *NRR 2024*, it has been proposed that noise monitoring shall be carried out on the first day of the updated lagoon show (i.e. 6 July 2024), as well as by the end of July, August and September 2024. The necessity and frequency of further noise monitoring should be further reviewed afterwards.

2.3 MONITORING LOCATIONS

Noise monitoring was conducted at five monitoring stations. These monitoring stations are identical to those adopted in the monthly noise monitoring report for the lagoon show (i.e. *SOTO* and *VHK*) of Ocean Park which launched on and off between 2020 and 2023. The locations of the five monitoring stations are indicated in **Figure 1** and presented in **Table 1** below.

TABLE 1 NOISE MONITORING STATIONS

| Noise Monitoring Station | Description | Location | With or without Façade Correction |
|--------------------------|--|---|-----------------------------------|
| AON1 | Open area adjacent to Police Training School | 1.2m above street level | Without façade correction |
| AON2 | Marriott Hotel, Ocean Park | 1m from façade at roof level | With façade correction |
| AON3 | Woodgreen Estate | 1.5m above street level outside boundary wall | With façade correction |
| AON4 | Manly Villa | 1.2 above street level | With façade correction |
| AON5 | Hau Yuen | 3.0m above street level outside boundary wall | With façade correction |

2.4 MONITORING PARAMETERS

NOISE LEVEL OF SHOW

Measurements of $L_{Aeq(5min)}$ readings were carried out to monitoring the noise level during the show. The schedule of show in September 2024 is shown in **Table 2**.




Figure 1
Noise Monitoring Locations



TABLE 2 SCHEDULE OF LAGOON NIGHT SHOW IN SEPTEMBER 2024

| Description | Time |
|---------------------------|---|
| Vision of Hong Kong (VHK) | <ul style="list-style-type: none"> Once between 1900 and 2200 hrs on each Saturday and Sunday, and once between 1900 and 2200 hrs on 18 September 2024 Lasted for around 5 minutes |
| Soul of the Ocean (SOTO) | <ul style="list-style-type: none"> Once between 1900 and 2200 hrs on each Saturday and Sunday, and once between 1900 and 2200 hrs on 18 September 2024 Lasted for around 12 minutes |

BACKGROUND NOISE LEVEL

Measurement of $L_{Aeq(5min)}$ were carried out before and after the show when speakers were switched off to calculate the overall background noise level at each location.

Any significant influencing factors on the measured noise levels were noted in accordance with standard acoustical principles and practices. The background-corrected noise level due to the show was computed based on the background noise level and measured noise level during the shows.

2.5 MONITORING FREQUENCY

Noise monitoring for the show has been conducted once in September 2024. The noise monitoring schedule is summarised in **Table 3**.

TABLE 3 NOISE MONITORING SCHEDULE IN SEPTEMBER 2024

| Monitoring Date | Monitoring Stations |
|------------------------------|---------------------|
| 28 September 2024 (Saturday) | AON1 to AON5 |

2.6 MONITORING METHODOLOGY

The sound level meters and calibrators used for the noise monitoring, as listed in **Table 4** below, complies with IEC 651: 1979 and 804:1985 (Type 1) or equivalent international standards.

TABLE 4 NOISE MONITORING EQUIPMENT

| Noise Monitoring Station | Monitoring Equipment |
|--------------------------|---|
| AON1 | Rion NL-52 Sound Level Meter, CAL200 Calibrator |
| AON2 | Rion NL-52 Sound Level Meter, CAL200 Calibrator |
| AON3 | Rion NL-52 Sound Level Meter, CAL200 Calibrator |
| AON4 | Rion NL-52 Sound Level Meter, CAL200 Calibrator |
| AON5 | Rion NL-52 Sound Level Meter, CAL200 Calibrator |

Noise monitoring was conducted with reference to the calibration and measurement procedures as stated in the *Technical Memorandum for the Assessment of Noise from Places other than Domestic Premises, Public Places or Construction Sites (IND-TM)* issued under the *Noise Control Ordinance (NCO)*. Immediately prior to and following each noise measurement the accuracy of the monitoring equipment was checked using an acoustic calibrator generating a known sound pressure level at a known frequency. Measurements were accepted as the calibration levels from before and after the noise measurement agree to within 1.0 dB.

The sound level meters and acoustic calibrators have been calibrated by a HOKLAS accredited laboratory every two years. The relevant calibration certificates are presented in **Appendix A**.

Noise measurements were conducted without the presence of fog and rain, and with steady wind speed and gusts not exceeding 5 ms⁻¹ and 10 ms⁻¹, respectively in accordance with international standards and practices (i.e. ISO 11819-1:1997 and ISO/FDIS 13472-1:2001). Measurement of L_{Aeq}, L₁₀, L₉₀, L_{max} and L_{min} has been recorded for reference.

If measured noise level is affected by other noise sources at the monitoring station, e.g. traffic noise, such that the measured noise level is dominated by noise source other than the show, noise data will be discarded.

If measured noise level for the show is below or equal to the measured background noise level, the noise from the show is considered as insignificant and hence negligible at the monitoring location.

2.7 COMPLIANCE ASSESSMENT

FIXED PLANT NOISE CRITERIA

As recommended in the approved EIA Report and stated in the EM&A Manual, OPC will follow the Action and Limit (A/L) Levels as recommended in the approved EIA Report and EM&A Manual which are summarised in **Table 5**. In case exceedances are resulted from cumulative impacts, all step stipulated in the Event/Action Plan shall be followed.

TABLE 5 ACTION AND LIMIT LEVELS FOR ENTERTAINMENT NOISE

| Noise Monitoring Station | Action Level | Limit Level |
|--------------------------|---|---------------------------------|
| AON1 | When documented complaint is received from any of the sensitive receivers | L _{eq} (5min) 60 dB(A) |
| AON2 | | L _{eq} (5min) 60 dB(A) |
| AON3 | | L _{eq} (5min) 55 dB(A) |
| AON4 | | L _{eq} (5min) 55 dB(A) |
| AON5 | | L _{eq} (5min) 55 dB(A) |

NOISE CRITERIA FOR OUTDOOR ACTIVITIES

The same set of noise criteria for outdoor activities as per the monthly noise monitoring reports prepared in 2020 to 2023 was adopted. The noise levels from the outdoor activities

should not be more than 5dB(A) above the prevailing background noise level during the daytime and evening periods (0700-2300 hrs), as measured at 1m from the exterior building façade of the most affected NSRs for regular outdoor activities. For the night-time period (2300-0700 hrs of the next day), the noise from outdoor events should not be audible at the nearby NSRs.

2.8 RESULTS OF NOISE MONITORING

The results of noise monitoring conducted on 28 September 2024 are given in **Appendix B**, with summary of compliance shown in **Table 6**. Photographs taken at the monitoring stations are shown in **Appendix C**.

TABLE 6 COMPLIANCE OF NOISE MONITORING

| Date and Show | Noise Monitoring Station | Compliance | | |
|--------------------------|--------------------------|------------|-------------|------------------|
| | | <BGL+5 | Limit Level | Not Applicable |
| 28 September 2024 (SOTO) | AON1 | Yes | Yes | - |
| | AON2 | Yes | Yes | - |
| | AON3 | Yes | Yes | - |
| | AON4 | Yes | Yes | - |
| | AON5 | - | - | N/A ^a |
| 28 September 2024 (VHK) | AON1 | Yes | Yes | - |
| | AON2 | Yes | Yes | - |
| | AON3 | - | - | N/A ^a |
| | AON4 | Yes | Yes | - |
| | AON5 | Yes | Yes | - |

Note:

^a Measured noise data was considered being affected by other significant noise sources, i.e., traffic and idling heavy vehicle

2.9 SUMMARY OF NOISE EXCEEDANCES

No record of noise exceedances during the show in September 2024.

3. CONCLUSION

This is the noise monitoring report which summarises the noise monitoring results during the lagoon night show in **September 2024**.

The noise monitoring was carried out at five designated monitoring stations.

No noise exceedance has been recorded.



APPENDIX A CALIBRATION CERTIFICATES OF THE
NOISE MEASUREMENT EQUIPMENT



Certificate of Calibration 校正證書

Certificate No. : C243578
證書編號

ITEM TESTED / 送檢項目 (Job No. / 序引編號 : IC24-0986)

Date of Receipt / 收件日期 : 4 June 2024

Description / 儀器名稱 : Sound Level Meter
Manufacturer / 製造商 : Rion
Model No. / 型號 : NL-52
Serial No. / 編號 : 00643049
Supplied By / 委託者 : Envirotech Services Co.
Room 712, 7/F, My Loft, 9 Hoi Wing Road, Tuen Mun,
New Territories, Hong Kong

TEST CONDITIONS / 測試條件

Temperature / 溫度 : $(23 \pm 2)^{\circ}\text{C}$
Line Voltage / 電壓 : ---

Relative Humidity / 相對濕度 : $(50 \pm 25)\%$

TEST SPECIFICATIONS / 測試規範

Calibration


DATE OF TEST / 測試日期 : 30 June 2024

TEST RESULTS / 測試結果

The results apply to the particular unit-under-test only.
The results do not exceed specified limits. (after adjustment)
These limits refer to manufacturer's published tolerances as requested by the customer.
The results are detailed in the subsequent page(s).

The test equipment used for calibration are traceable to National Standards via :
- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory
- Hottinger Brüel & Kjær Calibration Laboratory, Denmark
- Agilent Technologies / Keysight Technologies
- Fluke Everett Service Center, USA

Tested By : 
測試 : H T Wong
Assistant Engineer

Certified By : 
核證 : K C Lee
Engineer

Date of Issue : 3 July 2024
簽發日期

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。

Certificate of Calibration

校正證書

Certificate No. : C243578
證書編號

- The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 12 hours, and switched on to warm up for over 10 minutes before the commencement of the test.
- Self-calibration using the internal standard (After Adjustment) was performed before the test from 6.1.1.2 to 6.3.2.
- The results presented are the mean of 3 measurements at each calibration point.
- Test equipment :

| Equipment ID | Description | Certificate No. |
|--------------|-------------------------------------|-----------------|
| CL280 | 40 MHz Arbitrary Waveform Generator | C240212 |
| CL281 | Multifunction Acoustic Calibrator | CDK2302738 |

- Test procedure : MA101N.

- Results :

- 6.1 Sound Pressure Level

- 6.1.1 Reference Sound Pressure Level

- 6.1.1.1 Before Adjustment

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | * 91.3 | ± 1.1 |

* Out of IEC 61672 Class 1 Limit

- 6.1.1.2 After Adjustment

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | 94.0 | ± 1.1 |

- 6.1.2 Linearity

| UUT Setting | | | | Applied Value | | UUT Reading (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | 94.0 (Ref.) |
| | | | | 104.00 | | 104.1 |
| | | | | 114.00 | | 114.2 |

IEC 61672 Class 1 Limit : ± 0.6 dB per 10 dB step and ± 1.1 dB for overall different.

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。

Certificate of Calibration

校正證書

Certificate No. : C243578

證書編號

6.2 Time Weighting

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | 94.0 | Ref. |
| | | | Slow | | | 94.0 | ± 0.3 |

6.3 Frequency Weighting

6.3.1 A-Weighting


| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|--------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 63 Hz | 67.7 | -26.2 ± 1.5 |
| | | | | | 125 Hz | 77.8 | -16.1 ± 1.5 |
| | | | | | 250 Hz | 85.3 | -8.6 ± 1.4 |
| | | | | | 500 Hz | 90.8 | -3.2 ± 1.4 |
| | | | | | 1 kHz | 94.0 | Ref. |
| | | | | | 2 kHz | 95.2 | +1.2 ± 1.6 |
| | | | | | 4 kHz | 95.0 | +1.0 ± 1.6 |
| | | | | | 8 kHz | 93.0 | -1.1 (+2.1 ; -3.1) |
| | | | | | 16 kHz | 86.0 | -6.6 (+3.5 ; -17.0) |

6.3.2 C-Weighting

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|--------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. | | |
| 30 - 130 | L _C | C | Fast | 94.00 | 63 Hz | 93.2 | -0.8 ± 1.5 |
| | | | | | 125 Hz | 93.8 | -0.2 ± 1.5 |
| | | | | | 250 Hz | 94.0 | 0.0 ± 1.4 |
| | | | | | 500 Hz | 94.0 | 0.0 ± 1.4 |
| | | | | | 1 kHz | 94.0 | Ref. |
| | | | | | 2 kHz | 93.8 | -0.2 ± 1.6 |
| | | | | | 4 kHz | 93.2 | -0.8 ± 1.6 |
| | | | | | 8 kHz | 91.1 | -3.0 (+2.1 ; -3.1) |
| | | | | | 16 kHz | 84.1 | -8.5 (+3.5 ; -17.0) |

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。



Certificate of Calibration

校正證書

Certificate No. : C243578
證書編號

Remarks : - UUT Microphone Model No. : UC-59 & S/N : 10446

- Mfr's Limit : IEC 61672 Class 1

- Uncertainties of Applied Value :

| | | |
|--------|------------------|--------------------------|
| 94 dB | : 63 Hz - 125 Hz | : ± 0.35 dB |
| | 250 Hz - 500 Hz | : ± 0.30 dB |
| | 1 kHz | : ± 0.20 dB |
| | 2 kHz - 4 kHz | : ± 0.35 dB |
| | 8 kHz | : ± 0.45 dB |
| | 16 kHz | : ± 0.70 dB |
| 104 dB | : 1 kHz | : ± 0.10 dB (Ref. 94 dB) |
| 114 dB | : 1 kHz | : ± 0.10 dB (Ref. 94 dB) |

- The uncertainties are for a confidence probability of not less than 95 %.

Note :

Only the original copy or the laboratory's certified true copy is valid.

The values given in this Certificate only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environment changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Sun Creation Engineering Limited shall not be liable for any loss or damage resulting from the use of the equipment.

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗所書面批准。



輝創工程有限公司

Sun Creation Engineering Limited

Calibration & Testing Laboratory

Certificate of Calibration

校正證書

Certificate No. : C240423

證書編號

ITEM TESTED / 送檢項目 (Job No. / 序引編號 : IC24-0020)

Date of Receipt / 收件日期 : 5 January 2024

Description / 儀器名稱 : Precision Acoustic Calibrator

Manufacturer / 製造商 : LARSON DAVIS

Model No. / 型號 : CAL200

Serial No. / 編號 : 16172

Supplied By / 委託者 : Envirotech Services Co.

Room 712, 7/F, My Loft, 9 Hoi Wing Road, Tuen Mun,
New Territories, Hong Kong

TEST CONDITIONS / 測試條件

Temperature / 溫度 : $(23 \pm 2)^{\circ}\text{C}$

Relative Humidity / 相對濕度 : $(50 \pm 25)\%$

Line Voltage / 電壓 : ---

TEST SPECIFICATIONS / 測試規範

Calibration check

DATE OF TEST / 測試日期 : 24 January 2024

TEST RESULTS / 測試結果

The results apply to the particular unit-under-test only.

The results do not exceed specified limits.

These limits refer to manufacturer's published tolerances as requested by the customer.


The results are detailed in the subsequent page(s).

The test equipment used for calibration are traceable to National Standards via :

- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory
- Hottinger Brüel & Kjær Calibration Laboratory, Denmark
- Agilent Technologies / Keysight Technologies
- Fluke Everett Service Center, USA


Tested By

測試


K C Lee
Engineer

Certified By

核證


H C Chan
Engineer

Date of Issue

簽發日期

24 January 2024

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。

Sun Creation Engineering Limited - Calibration & Testing Laboratory

c/o 4/F, 1 Hing On Lane, Tuen Mun, New Territories, Hong Kong

輝創工程有限公司 - 校正及檢測實驗室

c/o 香港新界屯門興安里一號四樓

Tel/電話: (852) 2927 2606

Fax/傳真: (852) 2744 8986

E-mail/電郵: callab@suncreation.com

Website/網址: www.suncreation.com

Certificate of Calibration

校正證書

Certificate No. : C240423
證書編號

- The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 12 hours before the commencement of the test.
- The results presented are the mean of 3 measurements at each calibration point.
- Test equipment :

| <u>Equipment ID</u> | <u>Description</u> | <u>Certificate No.</u> |
|---------------------|-----------------------------------|------------------------|
| CL130 | Universal Counter | C233799 |
| CL281 | Multifunction Acoustic Calibrator | CDK2302738 |
| TST150A | Measuring Amplifier | C221750 |

- Test procedure : MA100N.

- Results :

5.1 Sound Level Accuracy

| UUT Nominal Value | Measured Value (dB) | Mfr's Limit (dB) | Uncertainty of Measured Value (dB) |
|----------------------|------------------------|---------------------|---------------------------------------|
| 94 dB, 1 kHz | 93.90 | ± 0.2 | ± 0.20 |
| 114 dB, 1 kHz | 113.90 | | |

5.2 Frequency Accuracy

| UUT Nominal Value (kHz) | Measured Value (kHz) | Mfr's Limit | Uncertainty of Measured Value (Hz) |
|----------------------------|-------------------------|----------------|---------------------------------------|
| 1 | 1.000 | 1 kHz ± 1 % | ± 1 |

Remark : The uncertainties are for a confidence probability of not less than 95 %.

Note :

Only the original copy or the laboratory's certified true copy is valid.

The values given in this Certificate only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environment changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Sun Creation Engineering Limited shall not be liable for any loss or damage resulting from the use of the equipment.

Certificate of Calibration

校正證書

Certificate No. : C237486

證書編號

ITEM TESTED / 送檢項目 (Job No. / 序引編號 : IC23-2475) Date of Receipt / 收件日期 : 8 December 2023

Description / 儀器名稱 : Sound Level Meter

Manufacturer / 製造商 : Rion

Model No. / 型號 : NL-52

Serial No. / 編號 : 01010406

Supplied By / 委託者 : Envirotech Services Co.

Room 712, 7/F, My Loft, 9 Hoi Wing Road, Tuen Mun,
New Territories, Hong Kong

TEST CONDITIONS / 測試條件

Temperature / 溫度 : (23 ± 2)°C

Relative Humidity / 相對濕度 : (50 ± 25)%

Line Voltage / 電壓 : ---

TEST SPECIFICATIONS / 測試規範

Calibration check

DATE OF TEST / 測試日期 : 31 December 2023

TEST RESULTS / 測試結果

The results apply to the particular unit-under-test only.

The results do not exceed specified limits.

These limits refer to manufacturer's published tolerances as requested by the customer.

The results are detailed in the subsequent page(s).

The test equipment used for calibration are traceable to National Standards via :

- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory
- Hottinger Brüel & Kjær Calibration Laboratory, Denmark
- Agilent Technologies / Keysight Technologies
- Fluke Everett Service Center, USA

Tested By

測試

: 

H T Wong

Assistant Engineer

Certified By

核證

: 

K C Lee

Engineer

Date of Issue

簽發日期

: 3 January 2024

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。

Certificate of Calibration

校正證書

Certificate No. : C237486

證書編號

1. The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 12 hours, and switched on to warm up for over 10 minutes before the commencement of the test.
2. Self-calibration was performed before the test.
3. The results presented are the mean of 3 measurements at each calibration point.
4. Test equipment :

| Equipment ID | Description | Certificate No. |
|--------------|-------------------------------------|-----------------|
| CL280 | 40 MHz Arbitrary Waveform Generator | C230306 |
| CL281 | Multifunction Acoustic Calibrator | CDK2302738 |

5. Test procedure : MA101N.

6. Results :

- 6.1 Sound Pressure Level

- 6.1.1 Reference Sound Pressure Level

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | 94.0 | ± 1.1 |

- 6.1.2 Linearity

| UUT Setting | | | | Applied Value | | UUT Reading (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | 94.0 (Ref.) |
| | | | | 104.00 | | 104.1 |
| | | | | 114.00 | | 114.0 |

IEC 61672 Class 1 Limit : ± 0.6 dB per 10 dB step and ± 1.1 dB for overall different.

- 6.2 Time Weighting

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | 94.0 | Ref. |
| | | | Slow | | | 94.0 | ± 0.3 |

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。

Certificate of Calibration

校正證書

Certificate No. : C237486

證書編號

6.3 Frequency Weighting

6.3.1 A-Weighting

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|--------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 63 Hz | 67.7 | -26.2 ± 1.5 |
| | | | | | 125 Hz | 77.8 | -16.1 ± 1.5 |
| | | | | | 250 Hz | 85.3 | -8.6 ± 1.4 |
| | | | | | 500 Hz | 90.7 | -3.2 ± 1.4 |
| | | | | | 1 kHz | 94.0 | Ref. |
| | | | | | 2 kHz | 95.2 | +1.2 ± 1.6 |
| | | | | | 4 kHz | 95.0 | +1.0 ± 1.6 |
| | | | | | 8 kHz | 93.0 | -1.1 (+2.1 ; -3.1) |
| | | | | | 16 kHz | 86.0 | -6.6 (+3.5 ; -17.0) |

6.3.2 C-Weighting

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|--------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. | | |
| 30 - 130 | L _C | C | Fast | 94.00 | 63 Hz | 93.1 | -0.8 ± 1.5 |
| | | | | | 125 Hz | 93.8 | -0.2 ± 1.5 |
| | | | | | 250 Hz | 94.0 | 0.0 ± 1.4 |
| | | | | | 500 Hz | 94.0 | 0.0 ± 1.4 |
| | | | | | 1 kHz | 94.0 | Ref. |
| | | | | | 2 kHz | 93.8 | -0.2 ± 1.6 |
| | | | | | 4 kHz | 93.2 | -0.8 ± 1.6 |
| | | | | | 8 kHz | 91.1 | -3.0 (+2.1 ; -3.1) |
| | | | | | 16 kHz | 84.1 | -8.5 (+3.5 ; -17.0) |

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。



Certificate of Calibration

校正證書

Certificate No. : C237486
證書編號

Remarks : - UUT Microphone Model No. : UC-59 & S/N : 13748

- Mfr's Limit : IEC 61672 Class 1

- Uncertainties of Applied Value :

| | | |
|--------|------------------|--------------------------|
| 94 dB | : 63 Hz - 125 Hz | : ± 0.35 dB |
| | 250 Hz - 500 Hz | : ± 0.30 dB |
| | 1 kHz | : ± 0.20 dB |
| | 2 kHz - 4 kHz | : ± 0.35 dB |
| | 8 kHz | : ± 0.45 dB |
| | 16 kHz | : ± 0.70 dB |
| 104 dB | : 1 kHz | : ± 0.10 dB (Ref. 94 dB) |
| 114 dB | : 1 kHz | : ± 0.10 dB (Ref. 94 dB) |

- The uncertainties are for a confidence probability of not less than 95 %.

Note :

Only the original copy or the laboratory's certified true copy is valid.

The values given in this Certificate only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environment changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Sun Creation Engineering Limited shall not be liable for any loss or damage resulting from the use of the equipment.

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室所書面批准。

Certificate of Calibration

for

Description: *Sound Level Calibrator*
Manufacturer: *Larson Davis*
Type No.: *CAL200*
Serial No.: *11333*

Submitted by:

Customer: *Envirotech Services Co.*
Address: *Rm.712, 7/F., My Loft, 9 Hoi Wing Road,
Tuen Mun, Hong Kong*

Upon receipt for calibration, the instrument was found to be:

Within

Outside

the allowable tolerance.

The test equipments used for calibration are traceable to National Standards via:

- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory

Date of receipt: 24 July 2024

Date of calibration: 26 July 2024

Date of NEXT calibration: 25 July 2025

Calibrated by: *David*
Calibration Technician

Certified by: *Ng Yan Wa*
*Mr. Ng Yan Wa
Laboratory Manager*

Date of issue: 26 July 2024



Certificate No.: APJ24-045-CC002

Page 1 of 2

1. Calibration Precautions:

- The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 24 hours, and switched on to warm up for over 10 minutes before the commencement of the test.
- The results presented are the mean of 3 measurements at each calibration point.

2. Calibration Specifications:

Calibration check

3. Calibration Conditions:

Air Temperature: 24.3 °C
 Air Pressure: 1004 hPa
 Relative Humidity: 57.9 %

4. Calibration Equipment:

| Test Equipment | Type | Serial No. | Calibration Report Number | Traceable to |
|--------------------------|------------|------------|---------------------------|--------------|
| Multifunction Calibrator | B&K 4226 | 2288467 | AV240081 | HOKLAS |
| Sound Level Meter | RION NA-28 | 30721812 | AV230128 | HOKLAS |

5. Calibration Results

5.1 Sound Pressure Level

| Nominal value dB | Accept lower level dB | Accept upper level dB | Measured value dB |
|---------------------|--------------------------|--------------------------|----------------------|
| 94.0 | 93.6 | 94.4 | 93.6 |

Note:

The values given in this certification only related to the values measured at the time of the calibration.



Certificate of Calibration

校正證書

Certificate No. : C240424
證書編號

ITEM TESTED / 送檢項目 (Job No. / 序引編號 : IC24-0020)

Date of Receipt / 收件日期 : 5 January 2024

Description / 儀器名稱 : Sound Level Meter

Manufacturer / 製造商 : Rion

Model No. / 型號 : NL-52

Serial No. / 編號 : 00710259

Supplied By / 委託者 : Envirotech Services Co.

Room 712, 7/F, My Loft, 9 Hoi Wing Road, Tuen Mun,
New Territories, Hong Kong

TEST CONDITIONS / 測試條件

Temperature / 溫度 : $(23 \pm 2)^{\circ}\text{C}$

Relative Humidity / 相對濕度 : $(50 \pm 25)\%$

Line Voltage / 電壓 : ---

TEST SPECIFICATIONS / 測試規範

Calibration check

DATE OF TEST / 測試日期 : 24 January 2024

TEST RESULTS / 測試結果

The results apply to the particular unit-under-test only.

The results do not exceed specified limits.

These limits refer to manufacturer's published tolerances as requested by the customer.

The results are detailed in the subsequent page(s).


The test equipment used for calibration are traceable to National Standards via :

- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory
- Hottinger Brüel & Kjær Calibration Laboratory, Denmark
- Agilent Technologies / Keysight Technologies
- Fluke Everett Service Center, USA

Tested By

測試


:


K C Lee
Engineer

Certified By

核證

:


H C Chan
Engineer

Date of Issue

簽發日期

:

24 January 2024

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。

Certificate of Calibration

校正證書

Certificate No. : C240424
證書編號

- The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 12 hours, and switched on to warm up for over 10 minutes before the commencement of the test.
- Self-calibration was performed before the test.
- The results presented are the mean of 3 measurements at each calibration point.
- Test equipment :

| Equipment ID | Description | Certificate No. |
|--------------|-------------------------------------|-----------------|
| CL280 | 40 MHz Arbitrary Waveform Generator | C240212 |
| CL281 | Multifunction Acoustic Calibrator | CDK2302738 |

- Test procedure : MA101N.

- Results :

- 6.1 Sound Pressure Level

- 6.1.1 Reference Sound Pressure Level

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | 94.0 | ± 1.1 |

- 6.1.2 Linearity

| UUT Setting | | | | Applied Value | | UUT Reading (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | 94.0 (Ref.) |
| | | | | 104.00 | | 104.0 |
| | | | | 114.00 | | 114.0 |

IEC 61672 Class 1 Limit : ± 0.6 dB per 10 dB step and ± 1.1 dB for overall different.

- 6.2 Time Weighting

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | 94.0 | Ref. |
| | | | Slow | | | 94.0 | ± 0.3 |

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。

Certificate of Calibration

校正證書

Certificate No. : C240424

證書編號

6.3 Frequency Weighting

6.3.1 A-Weighting

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|--------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 63 Hz | 67.7 | -26.2 ± 1.5 |
| | | | | | 125 Hz | 77.8 | -16.1 ± 1.5 |
| | | | | | 250 Hz | 85.3 | -8.6 ± 1.4 |
| | | | | | 500 Hz | 90.7 | -3.2 ± 1.4 |
| | | | | | 1 kHz | 94.0 | Ref. |
| | | | | | 2 kHz | 95.2 | +1.2 ± 1.6 |
| | | | | | 4 kHz | 95.0 | +1.0 ± 1.6 |
| | | | | | 8 kHz | 92.9 | -1.1 (+2.1 ; -3.1) |
| | | | | | 16 kHz | 86.0 | -6.6 (+3.5 ; -17.0) |

6.3.2 C-Weighting

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|--------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. | | |
| 30 - 130 | L _C | C | Fast | 94.00 | 63 Hz | 93.1 | -0.8 ± 1.5 |
| | | | | | 125 Hz | 93.8 | -0.2 ± 1.5 |
| | | | | | 250 Hz | 94.0 | 0.0 ± 1.4 |
| | | | | | 500 Hz | 94.0 | 0.0 ± 1.4 |
| | | | | | 1 kHz | 94.0 | Ref. |
| | | | | | 2 kHz | 93.8 | -0.2 ± 1.6 |
| | | | | | 4 kHz | 93.2 | -0.8 ± 1.6 |
| | | | | | 8 kHz | 91.0 | -3.0 (+2.1 ; -3.1) |
| | | | | | 16 kHz | 84.1 | -8.5 (+3.5 ; -17.0) |

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室所書面批准。

Certificate of Calibration

校正證書

Certificate No. : C240424
證書編號

Remarks : - UUT Microphone Model No. : UC-59 & S/N : 12128

- Mfr's Limit : IEC 61672 Class 1

- Uncertainties of Applied Value :

| | | |
|--------|------------------|--------------------------|
| 94 dB | : 63 Hz - 125 Hz | : ± 0.35 dB |
| | 250 Hz - 500 Hz | : ± 0.30 dB |
| | 1 kHz | : ± 0.20 dB |
| | 2 kHz - 4 kHz | : ± 0.35 dB |
| | 8 kHz | : ± 0.45 dB |
| | 16 kHz | : ± 0.70 dB |
| 104 dB | : 1 kHz | : ± 0.10 dB (Ref. 94 dB) |
| 114 dB | : 1 kHz | : ± 0.10 dB (Ref. 94 dB) |

- The uncertainties are for a confidence probability of not less than 95 %.

Note :

Only the original copy or the laboratory's certified true copy is valid.

The values given in this Certificate only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environment changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Sun Creation Engineering Limited shall not be liable for any loss or damage resulting from the use of the equipment.

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。

Certificate of Calibration

校正證書

Certificate No. : C237485

證書編號

ITEM TESTED / 送檢項目 (Job No. / 序引編號 : IC23-2475) Date of Receipt / 收件日期 : 8 December 2023

Description / 儀器名稱 : Precision Acoustic Calibrator

Manufacturer / 製造商 : LARSON DAVIS

Model No. / 型號 : CAL200

Serial No. / 編號 : 15678

Supplied By / 委託者 : Envirotech Services Co.

Room 712, 7/F, My Loft, 9 Hoi Wing Road, Tuen Mun,
New Territories, Hong Kong

TEST CONDITIONS / 測試條件

Temperature / 溫度 : $(23 \pm 2)^{\circ}\text{C}$

Relative Humidity / 相對濕度 : $(50 \pm 25)\%$

Line Voltage / 電壓 : ---

TEST SPECIFICATIONS / 測試規範

Calibration check

DATE OF TEST / 測試日期 : 31 December 2023

TEST RESULTS / 測試結果

The results apply to the particular unit-under-test only.

The results do not exceed specified limits.

These limits refer to manufacturer's published tolerances as requested by the customer.

The results are detailed in the subsequent page(s).

The test equipment used for calibration are traceable to National Standards via :

- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory
- Hottinger Brüel & Kjær Calibration Laboratory, Denmark
- Agilent Technologies / Keysight Technologies
- Fluke Everett Service Center, USA

Tested By

測試

:


H T Wong
Assistant Engineer

Certified By

核證

:


K C Lee
Engineer

Date of Issue

簽發日期

:

3 January 2024

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。

Certificate of Calibration

校正證書

Certificate No. : C237485

證書編號

- The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 12 hours before the commencement of the test.
- The results presented are the mean of 3 measurements at each calibration point.
- Test equipment :

| <u>Equipment ID</u> | <u>Description</u> | <u>Certificate No.</u> |
|---------------------|-----------------------------------|------------------------|
| CL130 | Universal Counter | C233799 |
| CL281 | Multifunction Acoustic Calibrator | CDK2302738 |
| TST150A | Measuring Amplifier | C221750 |

- Test procedure : MA100N.

- Results :

5.1 Sound Level Accuracy

| UUT Nominal Value | Measured Value (dB) | Mfr's Limit (dB) | Uncertainty of Measured Value (dB) |
|----------------------|------------------------|---------------------|---------------------------------------|
| 94 dB, 1 kHz | 93.90 | ± 0.2 | ± 0.20 |
| 114 dB, 1 kHz | 113.90 | | |

5.2 Frequency Accuracy

| UUT Nominal Value (kHz) | Measured Value (kHz) | Mfr's Limit | Uncertainty of Measured Value (Hz) |
|----------------------------|-------------------------|----------------|---------------------------------------|
| 1 | 1.000 | 1 kHz ± 1 % | ± 1 |

Remark : The uncertainties are for a confidence probability of not less than 95 %.

Note :

Only the original copy or the laboratory's certified true copy is valid.

The values given in this Certificate only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environment changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Sun Creation Engineering Limited shall not be liable for any loss or damage resulting from the use of the equipment.



Certificate of Calibration 校正證書

Certificate No. : C237046
證書編號

ITEM TESTED / 送檢項目 (Job No. / 序引編號 : IC23-2316) Date of Receipt / 收件日期 : 15 November 2023

Description / 儀器名稱 : Sound Level Meter
Manufacturer / 製造商 : Rion
Model No. / 型號 : NL-52
Serial No. / 編號 : 00175561
Supplied By / 委託者 : Envirotech Services Co.
Room 712, 7/F, My Loft, 9 Hoi Wing Road, Tuen Mun,
New Territories, Hong Kong

TEST CONDITIONS / 測試條件

Temperature / 溫度 : $(23 \pm 2)^{\circ}\text{C}$ Relative Humidity / 相對濕度 : $(50 \pm 25)\%$
Line Voltage / 電壓 : ---

TEST SPECIFICATIONS / 測試規範

Calibration check

DATE OF TEST / 測試日期 : 6 December 2023

TEST RESULTS / 測試結果

The results apply to the particular unit-under-test only.
The results do not exceed specified limits.
These limits refer to manufacturer's published tolerances as requested by the customer.
The results are detailed in the subsequent page(s).

The test equipment used for calibration are traceable to National Standards via :
- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory
- Hottinger Brüel & Kjær Calibration Laboratory, Denmark
- Agilent Technologies / Keysight Technologies
- Fluke Everett Service Center, USA

Tested By : 
測試 : _____
C K Lo
Project Engineer

Certified By : 
核證 : _____
K Q Lee
Engineer

Date of Issue : 6 December 2023
簽發日期

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗所書面批准。

Certificate of Calibration

校正證書

Certificate No. : C237046
證書編號

- The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 12 hours, and switched on to warm up for over 10 minutes before the commencement of the test.
- Self-calibration was performed before the test.
- The results presented are the mean of 3 measurements at each calibration point.
- Test equipment :

| Equipment ID | Description | Certificate No. |
|--------------|-------------------------------------|-----------------|
| CL280 | 40 MHz Arbitrary Waveform Generator | C230306 |
| CL281 | Multifunction Acoustic Calibrator | CDK2302738 |

5. Test procedure : MA101N.

6. Results :

6.1 Sound Pressure Level

6.1.1 Reference Sound Pressure Level

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | 93.2 | ± 1.1 |

6.1.2 Linearity

| UUT Setting | | | | Applied Value | | UUT Reading (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | 93.2 (Ref.) |
| | | | | 104.00 | | 103.3 |
| | | | | 114.00 | | 113.4 |

IEC 61672 Class 1 Limit : ± 0.6 dB per 10 dB step and ± 1.1 dB for overall different.

6.2 Time Weighting

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|-------------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. (kHz) | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 1 | 93.2 | Ref. |
| | | | Slow | | | 93.2 | ± 0.3 |

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗所書面批准。

Certificate of Calibration

校正證書

Certificate No. : C237046
證書編號

6.3 Frequency Weighting

6.3.1 A-Weighting

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|--------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. | | |
| 30 - 130 | L _A | A | Fast | 94.00 | 63 Hz | 66.9 | -26.2 ± 1.5 |
| | | | | | 125 Hz | 77.0 | -16.1 ± 1.5 |
| | | | | | 250 Hz | 84.5 | -8.6 ± 1.4 |
| | | | | | 500 Hz | 89.9 | -3.2 ± 1.4 |
| | | | | | 1 kHz | 93.2 | Ref. |
| | | | | | 2 kHz | 94.4 | +1.2 ± 1.6 |
| | | | | | 4 kHz | 94.2 | +1.0 ± 1.6 |
| | | | | | 8 kHz | 92.1 | -1.1 (+2.1 ; -3.1) |
| | | | | | 16 kHz | 85.2 | -6.6 (+3.5 ; -17.0) |

6.3.2 C-Weighting

| UUT Setting | | | | Applied Value | | UUT Reading (dB) | IEC 61672 Class 1 Limit (dB) |
|-------------|----------------|---------------------|----------------|---------------|--------|------------------|------------------------------|
| Range (dB) | Function | Frequency Weighting | Time Weighting | Level (dB) | Freq. | | |
| 30 - 130 | L _C | C | Fast | 94.00 | 63 Hz | 92.3 | -0.8 ± 1.5 |
| | | | | | 125 Hz | 93.0 | -0.2 ± 1.5 |
| | | | | | 250 Hz | 93.2 | 0.0 ± 1.4 |
| | | | | | 500 Hz | 93.2 | 0.0 ± 1.4 |
| | | | | | 1 kHz | 93.2 | Ref. |
| | | | | | 2 kHz | 93.0 | -0.2 ± 1.6 |
| | | | | | 4 kHz | 92.4 | -0.8 ± 1.6 |
| | | | | | 8 kHz | 90.2 | -3.0 (+2.1 ; -3.1) |
| | | | | | 16 kHz | 83.3 | -8.5 (+3.5 ; -17.0) |

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室所書面批准。



Certificate of Calibration

校正證書

Certificate No. : C237046
證書編號

Remarks : - UUT Microphone Model No. : UC-59 & S/N : 16651

- Mfr's Limit : IEC 61672 Class 1

- Uncertainties of Applied Value :

| | | |
|--------|------------------|--------------------------|
| 94 dB | : 63 Hz - 125 Hz | : ± 0.35 dB |
| | 250 Hz - 500 Hz | : ± 0.30 dB |
| | 1 kHz | : ± 0.20 dB |
| | 2 kHz - 4 kHz | : ± 0.35 dB |
| | 8 kHz | : ± 0.45 dB |
| | 16 kHz | : ± 0.70 dB |
| 104 dB | : 1 kHz | : ± 0.10 dB (Ref. 94 dB) |
| 114 dB | : 1 kHz | : ± 0.10 dB (Ref. 94 dB) |

- The uncertainties are for a confidence probability of not less than 95 %.

Note :

Only the original copy or the laboratory's certified true copy is valid.

The values given in this Certificate only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environment changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Sun Creation Engineering Limited shall not be liable for any loss or damage resulting from the use of the equipment.

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。



Certificate of Calibration

校正證書

Certificate No. : C242738
證書編號

ITEM TESTED / 送檢項目 (Job No. / 序引編號 : IC24-0781)

Date of Receipt / 收件日期 : 3 May 2024

Description / 儀器名稱 : Precision Acoustic Calibrator
Manufacturer / 製造商 : LARSON DAVIS
Model No. / 型號 : CAL200
Serial No. / 編號 : 11334
Supplied By / 委託者 : Envirotech Services Co.
Room 712, 7/F, My Loft, 9 Hoi Wing Road, Tuen Mun,
New Territories, Hong Kong

TEST CONDITIONS / 測試條件

Temperature / 溫度 : $(23 \pm 2)^{\circ}\text{C}$
Line Voltage / 電壓 : ---

Relative Humidity / 相對濕度 : $(50 \pm 25)\%$

TEST SPECIFICATIONS / 測試規範

Calibration check


DATE OF TEST / 測試日期 : 19 May 2024

TEST RESULTS / 測試結果

The results apply to the particular unit-under-test only.
The results do not exceed specified limits.
These limits refer to manufacturer's published or user's specified tolerances as requested by the customer.
The results are detailed in the subsequent page(s).

The test equipment used for calibration are traceable to National Standards via :
- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory
- Hottinger Brüel & Kjær Calibration Laboratory, Denmark
- Agilent Technologies / Keysight Technologies
- Fluke Everett Service Center, USA

Tested By : 
測試 : H T Wong
Assistant Engineer

Certified By : 
核證 : K C Lee
Engineer

Date of Issue : 20 May 2024
簽發日期

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.
本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。

Certificate of Calibration

校正證書

Certificate No. : C242738
證書編號

- The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 12 hours before the commencement of the test.
- The results presented are the mean of 3 measurements at each calibration point.
- Test equipment :

| <u>Equipment ID</u> | <u>Description</u> | <u>Certificate No.</u> |
|---------------------|-----------------------------------|------------------------|
| CL130 | Universal Counter | C233799 |
| CL281 | Multifunction Acoustic Calibrator | CDK2302738 |
| TST150A | Measuring Amplifier | C241879 |

- Test procedure : MA100N.

- Results :

5.1 Sound Level Accuracy

| UUT Nominal Value | Measured Value (dB) | User's Limit (dB) | Uncertainty of Measured Value (dB) |
|----------------------|------------------------|----------------------|---------------------------------------|
| 94 dB, 1 kHz | 93.60 | ± 0.5 | ± 0.20 |
| 114 dB, 1 kHz | 113.60 | | |

5.2 Frequency Accuracy

| UUT Nominal Value (kHz) | Measured Value (kHz) | Mfr's Limit | Uncertainty of Measured Value (Hz) |
|----------------------------|-------------------------|----------------|---------------------------------------|
| 1 | 1.000 | 1 kHz ± 1 % | ± 1 |

Remarks : - The user's limit is a customer pre-defined operating tolerance of the UUT, suitable for one's own intended use.

- The uncertainties are for a confidence probability of not less than 95 %.

Note :

Only the original copy or the laboratory's certified true copy is valid.

The values given in this Certificate only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environment changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Sun Creation Engineering Limited shall not be liable for any loss or damage resulting from the use of the equipment.

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。

Certificate of Calibration

for

Description: *Sound Level Meter*
Manufacturer: *RION*
Type No.: *NL-52 (Serial No.: 00542913)*
Microphone: *UC-53A (Serial No.: 99995)*
Preamplifier: *NH-25 (Serial No.:43068)*

Submitted by:

Customer: *Envirotech Services Co.*
Address: *Rm.712, 7/F., My Loft, 9 Hoi Wing Road,
Tuen Mun, Hong Kong*

Upon receipt for calibration, the instrument was found to be:

- Within (31.5Hz – 8kHz)
 Outside

the allowable tolerance.

The test equipment used for calibration are traceable to National Standards via:

- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory

Date of receipt: 28 August 2024

Date of calibration: 29 August 2024

Date of NEXT calibration: 28 August 2025

Calibrated by: _____
Calibration Technician

Certified by: _____
*Mr. Ng Yan Wa
Laboratory Manager*

Date of issue: 29 August 2024



Certificate No.: APJ24-058-CC001

Page 1 of 4

1. Calibration Precaution:

- The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 24 hours, and switched on to warm up for over 10 minutes before the commencement of the test.
- The results presented are the mean of 3 measurements at each calibration point.

2. Calibration Conditions:

Air Temperature: 24.6°C
 Air Pressure: 1004 hPa
 Relative Humidity: 53.9%

3. Calibration Equipment:

| | Type | Serial No. | Calibration Report Number | Traceable to |
|--------------------------|----------|------------|---------------------------|--------------|
| Multifunction Calibrator | B&K 4226 | 2288467 | AV240081 | HOKLAS |

4. Calibration Results

Sound Pressure Level

Reference Sound Pressure Level

| Setting of Unit-under-test (UUT) | | | | Applied value | | UUT Reading, | IEC 61672 Class 1 |
|----------------------------------|-----------------|----------------|-----------|---------------|------|-------------------|-------------------|
| Range, dB | Freq. Weighting | Time Weighting | Level, dB | Frequency, Hz | dB | Specification, dB | |
| 30-130 | dBa SPL | Fast | 94 | 1000 | 94.0 | ±0.4 | |

Linearity

| Setting of Unit-under-test (UUT) | | | | Applied value | | UUT Reading, | IEC 61672 Class 1 |
|----------------------------------|-----------------|----------------|-----------|---------------|-------|-------------------|-------------------|
| Range, dB | Freq. Weighting | Time Weighting | Level, dB | Frequency, Hz | dB | Specification, dB | |
| 30-130 | dBa SPL | Fast | 94 | 1000 | 94.0 | Ref | |
| | | | 104 | | 104.0 | ±0.3 | |
| | | | 114 | | 114.0 | ±0.3 | |

Time Weighting

| Setting of Unit-under-test (UUT) | | | | Applied value | | UUT Reading, | IEC 61672 Class 1 |
|----------------------------------|-----------------|----------------|-----------|---------------|------|-------------------|-------------------|
| Range, dB | Freq. Weighting | Time Weighting | Level, dB | Frequency, Hz | dB | Specification, dB | |
| 30-130 | dBa SPL | Fast | 94 | 1000 | 94.0 | Ref | |
| | | Slow | | | 94.0 | ±0.3 | |



Certificate No.: APJ24-058-CC001

Page 2 of 4

Frequency Response

Linear Response

| Setting of Unit-under-test (UUT) | | | Applied value | | UUT Reading, dB | IEC 61672 Class 1 Specification, dB | |
|----------------------------------|-----------------|----------------|---------------|---------------|--------------------|--|------|
| Range, dB | Freq. Weighting | Time Weighting | Level, dB | Frequency, Hz | | | |
| 30-130 | dB | SPL | Fast | 94 | 31.5 | 92.7 | ±2.0 |
| | | | | | 63 | 93.7 | ±1.5 |
| | | | | | 125 | 93.9 | ±1.5 |
| | | | | | 250 | 94.0 | ±1.4 |
| | | | | | 500 | 94.0 | ±1.4 |
| | | | | | 1000 | 94.0 | Ref |
| | | | | | 2000 | 93.9 | ±1.6 |
| | | | | | 4000 | 94.3 | ±1.6 |
| | | | | 8000 | 92.4 | +2.1; -3.1 | |

A-weighting

| Setting of Unit-under-test (UUT) | | | Applied value | | UUT Reading, dB | IEC 61672 Class 1 Specification, dB | |
|----------------------------------|-----------------|----------------|---------------|---------------|--------------------|--|-----------|
| Range, dB | Freq. Weighting | Time Weighting | Level, dB | Frequency, Hz | | | |
| 30-130 | dBA | SPL | Fast | 94 | 31.5 | 53.5 | -39.4±2.0 |
| | | | | | 63 | 67.5 | -26.2±1.5 |
| | | | | | 125 | 77.8 | -16.1±1.5 |
| | | | | | 250 | 85.3 | -8.6±1.4 |
| | | | | | 500 | 90.8 | -3.2±1.4 |
| | | | | | 1000 | 94.0 | Ref |
| | | | | | 2000 | 95.2 | +1.2±1.6 |
| | | | | | 4000 | 95.3 | +1.0±1.6 |
| | | | | 8000 | 91.3 | -1.1±2.1; -3.1 | |

C-weighting

| Setting of Unit-under-test (UUT) | | | Applied value | | UUT Reading, dB | IEC 61672 Class 1 Specification, dB | |
|----------------------------------|-----------------|----------------|---------------|---------------|--------------------|--|----------|
| Range, dB | Freq. Weighting | Time Weighting | Level, dB | Frequency, Hz | | | |
| 30-130 | dBC | SPL | Fast | 94 | 31.5 | 89.7 | -3.0±2.0 |
| | | | | | 63 | 92.9 | -0.8±1.5 |
| | | | | | 125 | 93.8 | -0.2±1.5 |
| | | | | | 250 | 94.0 | -0.0±1.4 |
| | | | | | 500 | 94.0 | -0.0±1.4 |
| | | | | | 1000 | 94.0 | Ref |
| | | | | | 2000 | 93.8 | -0.2±1.6 |
| | | | | | 4000 | 93.5 | -0.8±1.6 |
| | | | | 8000 | 89.4 | -3.0±2.1; -3.1 | |

Certificate No.: APJ24-058-CC001



Page 3 of 4

5. Calibration Results Applied

The results apply to the particular unit-under-test only. All calibration points are within manufacture's specification as IEC 61672 Class 1.

Uncertainties of Applied Value:

| | | |
|--------|---------|--------|
| 94 dB | 31.5 Hz | ± 0.10 |
| | 63 Hz | ± 0.10 |
| | 125 Hz | ± 0.10 |
| | 250 Hz | ± 0.10 |
| | 500 Hz | ± 0.10 |
| | 1000 Hz | ± 0.05 |
| | 2000 Hz | ± 0.05 |
| | 4000 Hz | ± 0.10 |
| | 8000 Hz | ± 0.10 |
| 104 dB | 1000 Hz | ± 0.05 |
| 114 dB | 1000 Hz | ± 0.05 |

The uncertainties are evaluated for a 95% confidence level.

Note:

The values given in this certification only related to the values measured at the time of the calibration and any uncertainties quoted will not allow for the equipment long-term drift, variations with environmental changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the calibration. (A+A)*L shall not be liable for any loss or damage resulting from the use of the equipment.



輝創工程有限公司

Sun Creation Engineering Limited

Calibration & Testing Laboratory

Certificate of Calibration

校正證書

Certificate No. : C240965

證書編號

ITEM TESTED / 送檢項目 (Job No. / 序引編號: IC24-0190)

Date of Receipt / 收件日期: 1 February 2024

Description / 儀器名稱 : Precision Acoustic Calibrator

Manufacturer / 製造商 : LARSON DAVIS

Model No. / 型號 : CAL200

Serial No. / 編號 : 10227

Supplied By / 委託者 : Envirotech Services Co.

Room 712, 7/F, My Loft, 9 Hoi Wing Road, Tuen Mun,
New Territories, Hong Kong

TEST CONDITIONS / 測試條件

Temperature / 溫度 : $(23 \pm 2)^{\circ}\text{C}$

Relative Humidity / 相對濕度 : $(50 \pm 25)\%$

Line Voltage / 電壓 : ---

TEST SPECIFICATIONS / 測試規範

Calibration check

DATE OF TEST / 測試日期 : 22 February 2024

TEST RESULTS / 測試結果

The results apply to the particular unit-under-test only.

The results are detailed in the subsequent page(s).

The test equipment used for calibration are traceable to National Standards via :

- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory
- Hottinger Brüel & Kjær Calibration Laboratory, Denmark
- Agilent Technologies / Keysight Technologies
- Fluke Everett Service Center, USA

Tested By
測試


K C Lee
Engineer

Certified By
核證


H C Chan
Engineer

Date of Issue
簽發日期

22 February 2024

The test equipment used for calibration is traceable to the National Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory.

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗室書面批准。

Certificate of Calibration

校正證書

Certificate No. : C240965

證書編號

- The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 12 hours before the commencement of the test.
- The results presented are the mean of 3 measurements at each calibration point.
- Test equipment :

| <u>Equipment ID</u> | <u>Description</u> | <u>Certificate No.</u> |
|---------------------|-----------------------------------|------------------------|
| CL130 | Universal Counter | C233799 |
| CL281 | Multifunction Acoustic Calibrator | CDK2302738 |
| TST150A | Measuring Amplifier | C221750 |

- Test procedure : MA100N.

- Results :

5.1 Sound Level Accuracy

| UUT Nominal Value | Measured Value (dB) | Uncertainty of Measured Value (dB) |
|----------------------|------------------------|---------------------------------------|
| 94 dB, 1 kHz | 93.90 | ± 0.20 |
| 114 dB, 1 kHz | 113.90 | |

5.2 Frequency Accuracy

| UUT Nominal Value (kHz) | Measured Value (kHz) | Uncertainty of Measured Value (Hz) |
|----------------------------|-------------------------|---------------------------------------|
| 1 | 1.000 | ± 1 |

Remark : The uncertainties are for a confidence probability of not less than 95 %.

Note :

Only the original copy or the laboratory's certified true copy is valid.

The values given in this Certificate only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environment changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Sun Creation Engineering Limited shall not be liable for any loss or damage resulting from the use of the equipment.



APPENDIX B RESULTS OF NOISE MONITORING

Noise Measurement Field Record Sheet

Project Name / GMS No.: 0540005 OPC Noise Monitoring for Updated Lagoon Show

Noise Monitoring Station: AON1

Noise Monitoring Staff: K. C. Ho

Noise Meter Model / Identification: Rion NL-52 / 00643049

Calibrator Model / Identification: CAL200 / 16172

| Date | Start | End Time | Monitoring Event | Leq, 5min | Leq, 15min | Average BGL | Noise Level with Show - BGL | BG-corrected Leq, 15min | NCO (BGL+5) | Limit Level | Compliance | | Remarks |
|-----------|-------|------------|------------------|-----------|------------|-------------|-----------------------------|-------------------------|-------------|-------------|-------------|-------------|---------|
| | | | | | | | | | | | NCO (BGL+5) | Limit Level | |
| 28-Sep-24 | 18:45 | 18:50 | Background | 68.3 | - | 65.3 | - | - | 70.3 | 60 | - | - | - |
| | 18:50 | 18:55 | Background | 66.0 | | | | | | | | | |
| | 18:55 | 19:00 | Background | 65.6 | | | | | | | | | |
| | 19:00 | 19:05 | SOTO | 61.8 | 62.4 | | -2.9 | Negligible | | | Yes | Yes | - |
| | 19:05 | 19:10 | SOTO | 63.5 | | | | | | | | | |
| | 19:10 | 19:15 | SOTO | 61.7 | 64.4 | | -0.9 | Negligible | | | Yes | Yes | - |
| | 19:15 | 19:20 | SOTO / VHK | 64.0 | | | | | | | | | |
| | 19:20 | 19:25 | VHK | 64.7 | - | | - | - | | | - | - | - |
| | 19:25 | 19:30 | Background | 64.4 | | | | | | | | | |
| | 19:30 | 19:35 | Background | 60.1 | | | | | | | | | |
| 19:35 | 19:40 | Background | 63.3 | | | | | | | | | | |

Schedule of Event:

28-9-2024: SOTO (19:02 - 19:15), VHK (19:16 - 19:22)

Note:

- (a) Average BGL was calculated using $L_{eq(5min)}$ 15 minutes before and after the show.
- (b) Impact from the show is considered negligible when the measured noise level with the show is equal to or lower than the measured background noise level.
- (c) The Schedule of Event was confirmed with OPC's staff on-site.
- (d) +3 dB(A) free-field correction has been applied to monitoring results at AON1.

Noise Measurement Field Record Sheet

Project Name / GMS No.: 0540005 OPC Noise Monitoring for Updated Lagoon Show

Noise Monitoring Station: AON2

Noise Monitoring Staff: Y. P. Fai

Noise Meter Model / Identification: Rion NL-52 / 01010406

Calibrator Model / Identification: CAL200 / 11333

| Date | Start | End Time | Monitoring Event | Leq, 5min | Leq, 15min | Average BGL | Noise Level with Show - BGL | BG-corrected Leq, 15min | NCO (BGL+5) | Limit Level | Compliance | | Remarks |
|-----------|-------|------------|------------------|-----------|------------|-------------|-----------------------------|-------------------------|-------------|-------------|-------------|-------------|---------|
| | | | | | | | | | | | NCO (BGL+5) | Limit Level | |
| 28-Sep-24 | 18:45 | 18:50 | Background | 64.4 | - | 63.2 | - | - | 68.2 | 60 | - | - | - |
| | 18:50 | 18:55 | Background | 64.3 | | | | | | | | | |
| | 18:55 | 19:00 | Background | 62.1 | | | | | | | | | |
| | 19:00 | 19:05 | SOTO | 63.7 | 64.0 | | 0.8 | 56.4 | | | Yes | Yes | - |
| | 19:05 | 19:10 | SOTO | 64.4 | | | | | | | | | |
| | 19:10 | 19:15 | SOTO | 64.0 | 64.9 | | 1.7 | 59.9 | | | Yes | Yes | - |
| | 19:15 | 19:20 | SOTO / VHK | 65.7 | | | | | | | | | |
| | 19:20 | 19:25 | VHK | 63.9 | | | | | | | | | |
| | 19:25 | 19:30 | Background | 62.2 | - | | - | - | | | - | - | - |
| | 19:30 | 19:35 | Background | 63.9 | | | | | | | | | |
| 19:35 | 19:40 | Background | 61.6 | | | | | | | | | | |

Schedule of Event:

28-9-2024: SOTO (19:02 - 19:15), VHK (19:16 - 19:22)

Note:

(a) Average BGL was calculated using $L_{eq(5min)}$ 15 minutes before and after the show.

(b) Impact from the show is considered negligible when the measured noise level with the show is equal to or lower than the measured background noise level.

(c) The Schedule of Event was confirmed with OPC's staff on-site.

Noise Measurement Field Record Sheet

Project Name / GMS No.: 0540005 OPC Noise Monitoring for Updated Lagoon Show

Noise Monitoring Station: AON3

Noise Monitoring Staff: Lap Kwok

Noise Meter Model / Identification: Rion NL-52 / 00710259

Calibrator Model / Identification: CAL200 / 15678

| Date | Start | End Time | Monitoring Event | Leq, 5min | Leq, 15min | Average BGL | Noise Level with Show - BGL | BG-corrected Leq, 15min | NCO (BGL+5) | Limit Level | Compliance | | Remarks | |
|-----------|-------|------------|------------------|-----------|------------|-------------|-----------------------------|-------------------------|-------------|-------------|-------------|-------------|----------------------------|---|
| | | | | | | | | | | | NCO (BGL+5) | Limit Level | | |
| 28-Sep-24 | 18:45 | 18:50 | Background | 64.5 | - | 63.4 | - | - | 68.4 | 55 | - | - | - | |
| | 18:50 | 18:55 | Background | 65.5 | | | - | - | | | - | | | |
| | 18:55 | 19:00 | Background | 61.9 | | | - | - | | | - | | | |
| | 19:00 | 19:05 | SOTO | 63.6 | 64.0 | | 0.6 | 54.9 | | | Yes | Yes | - | |
| | 19:05 | 19:10 | SOTO | 65.2 | | | 0.9 | 57.3 | | | N/A | N/A | Dominated by traffic noise | |
| | 19:10 | 19:15 | SOTO | 62.8 | 64.4 | | - | - | | | - | - | - | - |
| | 19:15 | 19:20 | SOTO / VHK | 64.2 | | | - | - | | | - | - | - | |
| | 19:20 | 19:25 | VHK | 64.5 | | | - | - | | | - | - | - | |
| | 19:25 | 19:30 | Background | 63.2 | - | | - | - | | | - | - | - | - |
| | 19:30 | 19:35 | Background | 62.1 | | | - | - | | | - | - | - | |
| 19:35 | 19:40 | Background | 61.8 | - | | - | - | - | - | | | | | |

Schedule of Event:

28-9-2024: SOTO (19:02 - 19:15), VHK (19:16 - 19:22)

Note:

(a) Average BGL was calculated using $L_{eq(5min)}$ 15 minutes before and after the show.

(b) Impact from the show is considered negligible when the measured noise level with the show is equal to or lower than the measured background noise level.

(c) The Schedule of Event was confirmed with OPC's staff on-site.

Noise Measurement Field Record Sheet

Project Name / GMS No.: 0540005 OPC Noise Monitoring for Updated Lagoon Show

Noise Monitoring Station: AON4

Noise Monitoring Staff: K. T. Ho

Noise Meter Model / Identification: Rion NL-52 / 00175561

Calibrator Model / Identification: CAL200 / 11334

| Date | Start | End Time | Monitoring Event | Leq, 5min | Leq, 15min | Average BGL | Noise Level with Show - BGL | BG-corrected Leq, 15min | NCO (BGL+5) | Limit Level | Compliance | | Remarks |
|-----------|-------|------------|------------------|-----------|------------|-------------|-----------------------------|-------------------------|-------------|-------------|-------------|-------------|---------|
| | | | | | | | | | | | NCO (BGL+5) | Limit Level | |
| 28-Sep-24 | 18:45 | 18:50 | Background | 55.8 | - | 55.6 | - | - | 60.6 | 55 | - | - | - |
| | 18:50 | 18:55 | Background | 56.9 | | | | | | | | | |
| | 18:55 | 19:00 | Background | 56.4 | | | | | | | | | |
| | 19:00 | 19:05 | SOTO | 55.2 | 56.1 | | 0.5 | 46.1 | | | Yes | Yes | - |
| | 19:05 | 19:10 | SOTO | 57.7 | | | | | | | | | |
| | 19:10 | 19:15 | SOTO | 54.8 | 56.9 | | 1.3 | 51.0 | | | Yes | Yes | - |
| | 19:15 | 19:20 | SOTO / VHK | 55.1 | | | | | | | | | |
| | 19:20 | 19:25 | VHK | 58.2 | - | | - | - | | | - | - | - |
| | 19:25 | 19:30 | Background | 54.5 | | | | | | | | | |
| | 19:30 | 19:35 | Background | 53.1 | | | | | | | | | |
| 19:35 | 19:40 | Background | 56.1 | | | | | | | | | | |

Schedule of Event:

28-9-2024: SOTO (19:02 - 19:15), VHK (19:16 - 19:22)

Note:

(a) Average BGL was calculated using $L_{eq(5min)}$ 15 minutes before and after the show.

(b) Impact from the show is considered negligible when the measured noise level with the show is equal to or lower than the measured background noise level.

(c) The Schedule of Event was confirmed with OPC's staff on-site.

Noise Measurement Field Record Sheet

Project Name / GMS No.: 0540005 OPC Noise Monitoring for Updated Lagoon Show

Noise Monitoring Station: AON5

Noise Monitoring Staff: Magnum Fan

Noise Meter Model / Identification: Rion NL-52 / 00542913

Calibrator Model / Identification: CAL200 / 10227

| Date | Start | End Time | Monitoring Event | Leq, 5min | Leq, 15min | Average BGL | Noise Level with Show - BGL | BG-corrected Leq, 15min | NCO (BGL+5) | Limit Level | Compliance | | Remarks |
|-----------|-------|------------|------------------|-----------|------------|-------------|-----------------------------|-------------------------|-------------|-------------|-------------|-------------|----------------------------|
| | | | | | | | | | | | NCO (BGL+5) | Limit Level | |
| 28-Sep-24 | 18:45 | 18:50 | Background | 58.4 | - | 57.0 | - | - | 62.0 | 55 | - | - | - |
| | 18:50 | 18:55 | Background | 57.5 | | | | | | | | | |
| | 18:55 | 19:00 | Background | 56.2 | | | | | | | | | |
| | 19:00 | 19:05 | SOTO | 62.1 | 60.9 | | 3.9 | 58.6 | | | N/A | N/A | Dominated by traffic noise |
| | 19:05 | 19:10 | SOTO | 58.6 | | | | | | | | | |
| | 19:10 | 19:15 | SOTO | 61.3 | 57.6 | | 0.6 | 48.5 | | | Yes | Yes | - |
| | 19:15 | 19:20 | SOTO / VHK | 58.1 | | | | | | | | | |
| | 19:20 | 19:25 | VHK | 57.0 | - | | - | - | | | - | - | - |
| | 19:25 | 19:30 | Background | 52.1 | | | | | | | | | |
| | 19:30 | 19:35 | Background | 56.3 | | | | | | | | | |
| 19:35 | 19:40 | Background | 58.8 | | | | | | | | | | |

Schedule of Event:

28-9-2024: SOTO (19:02 - 19:15), VHK (19:16 - 19:22)

Note:

(a) Average BGL was calculated using $L_{eq(5min)}$ 15 minutes before and after the show.

(b) Impact from the show is considered negligible when the measured noise level with the show is equal to or lower than the measured background noise level.

(c) The Schedule of Event was confirmed with OPC's staff on-site.



APPENDIX C PHOTOGRAPHS OF THE MONITORING STATIONS



Appendix C

Open Area adjacent to Police Training School (AON1)

DATE: 28/09/2024











Appendix C

Hau Yuen (AON5)

DATE: 28/09/2024





ERM HAS OVER 160 OFFICES ACROSS THE FOLLOWING COUNTRIES AND TERRITORIES WORLDWIDE

- | | |
|------------|-----------------|
| Argentina | The Netherlands |
| Australia | New Zealand |
| Belgium | Peru |
| Brazil | Poland |
| Canada | Portugal |
| China | Puerto Rico |
| Colombia | Romania |
| France | Senegal |
| Germany | Singapore |
| Ghana | South Africa |
| Guyana | South Korea |
| Hong Kong | Spain |
| India | Switzerland |
| Indonesia | Taiwan |
| Ireland | Tanzania |
| Italy | Thailand |
| Japan | UAE |
| Kazakhstan | UK |
| Kenya | US |
| Malaysia | Vietnam |
| Mexico | |
| Mozambique | |

ERM Hong Kong

2509, 25/F One Harbourfront
18 Tak Fung Street
Hung Hom, Kowloon
Hong Kong

T: +852 2271 3000

F: +852 3015 8052

www.erm.com