

Our ref.: LES/J2019-02/CS/L018

Date : 23 May 2019

Environmental Protection Department Environmental Assessment Division Metro Assessment Group (43) By Hand

27th floor, Southorn Centre, 130 Hennessy Road, Wan Chai, Hong Kong

Attn: Ms. Eva LAU (Env Protection Offr (Metro Assessment)43)

Dear Ms. LAU,

Contract No. DC/2018/05
Relocation of Sha Tin Sewage Treatment Works to Caverns – Site Preparation and Access Tunnel Construction

Proposal for Using Direct Reading Dust Meter for Construction Phase Air Monitoring

Submission under the Requirement of EM&A Manual

We refer to the comments from your email dated on 22 May 2019 regarding the captioned, we submit herewith 4 hard copies and 3 electronic copies of proposal with the following items for your approval.

- 1. Proposal for using direct reading dust meter for construction phase air monitoring;
- 2. Response to comments
- 3. Letter from the Independent Environmental Checker showing they have no further comment on this proposal.

Should you have any queries, please contact the undersigned at 9108 0531.

Yours faithfully, For and On Behalf Of

Lam Environmental Services Limited

Derek Lo

Environmental Team Leader

Encl.

c.c. DSD Mr. Kenneth Email w/e
AECOM (CRE Office) Mr. Simon Leung Email w/e
Mott MacDonald Hong Kong Mr. Brandon Wong Email w/e

Limited



Environmental Protection Department EIAO Register Office 27/F, Southorn Centre 130 Hennessy Road Wanchai, Hong Kong

Attn: Ms. Eva Lau

Your Reference

Our Reference EC/TC/BW/bw/T407129/ Correspondence/L016

3/F Mapletree Bay Point 348 Kwun Tong Road Kowloon Hong Kong

T +852 2828 5757 F +852 2827 1823 mottmac.hk **Sha Tin Cavern Sewage Treatment Works**

Environmental Permit No. EP-533/2017

Contract No. DC/2018/05 Relocation of Sha Tin Sewage Treatment Works to Caverns – Site Preparation and Access Tunnel Construction

<u>Proposal to use Direct Reading Dust Meter for Construction Phase 1-hour TSP</u> <u>Impact Air Quality Monitoring</u>

23 May 2019

By Hand

Dear Madam.

I refer to the letter dated 23 May 2019 (ref: LES/J2019-02/CS/L018) from the Environmental Team Leader regarding his proposal to use direct reading dust meter for construction phase 1-hour TSP impact air quality monitoring, pursuant to Sections 2.2.2.2 and 2.2.2.5 of the EM&A Manual for the captioned Project.

I have no comment on the captioned proposal and hereby verify it as conforming to the information and recommendations contained in the EM&A Manual, in accordance with Condition 3.1 of Environmental Permit No. EP-533/2017 and Section 2.2.1.5 of the EM&A Manual.

Should you have any queries regarding the captioned or require any further information, please contact the undersigned at 2828 5875.

Yours faithfully

for MOTT MACDONALD HONG KONG LIMITED

Brandon Wong

Independent Environmental Checker

T +852 2828 5875

Brandon.Wong@mottmac.com

Encl.

c.c. DSD

Mr. Kenneth Poon

By Email

AECOM Asia Co. Ltd.

Mr. Simon Leung

By Email

Lam Environmental Services Limited

Mr. Derek Lo

By Email



CONTRACT NO. SPW 25/2018

ENVIRONMENTAL TEAM FOR RELOCATION OF SHA TIN SEWAGE TREATMENT WORKS TO CAVERNS – SITE PREPARATION AND ACCESS TUNNEL CONSTRUCTION

UNDER ENVIRONMENTAL PERMIT NO. EP-533/2017

PROPOSAL FOR USING DIRECT READING DUST METER FOR CONSTRUCTION PHASE AIR MONITORING

CLIENTS:

Drainage Services Department

PREPARED BY:

Lam Environmental Services Limited

11/F Centre Point 181-185 Gloucester Road Wanchai, H.K.

Telephone: (852) 2882-3939
Facsimile: (852) 2882-3331
E-mail: info@lamenviro.com
Website: http://www.lamenviro.com

CERTIFIED BY:

Derek Lo

Environmental Team Leader

DATE:

23 May 2019



Table of Contents

1	Introduction	1
1.1	Background	1
1.2	Proposed Monitoring Equipment	1
1.3	Monitoring Methodology	. 3

LIST OF FIGURES

Figure 1.1 Air quality monitoring stations

LIST OF APPENDIXES

Appendix A Specification of equipment

Appendix B Sample of equipment calibration certification

Appendix C Project layout plan



1 Introduction

1.1 Background

- 1.1.1. Lam Environmental Services Limited has been commissioned by Drainage Services Department to undertake the Environmental Monitoring and Audit works for Contract No. SPW 25/2018 Environmental Team for Relocation of Sha Tin Sewage Treatment Works to Caverns Site Preparation and Access Tunnel Construction, i.e. Contract No. DC/2018/05.
- 1.1.2. According to EM&A Manual Section 2.2.1.5,1-hour TSP levels should be measured by following the standard method as set out in High Volume Sampling Method for Total Suspended Particulates, Part 50 Chapter 1 Appendix B, Title 40 of the Code of Federal Regulations of the US Environmental Protection Agency (USEPA) (hereinafter referred to as "High Volume Sampler (HVS) method"). Upon approval of EPD and IEC, an alternative sampling method of using direct reading methods which are capable of producing comparable results as that by the high volume sampling method can be used to indicate short event impacts.
- 1.1.3. This proposal comprised of the monitoring method proposed as per EM&A manual Section S2.2.1.5 for seeking approval of EPD and IEC, the alternative sampling method of using direct reading methods for which the monitoring result is comparable with the high volume sampling method can be used for the impact air monitoring.

1.2 Proposed Monitoring Equipment

- 1.2.1. With reference to S1.2.5 of Appendix D2 of the General Technical Requirements of Environmental Monitoring Guidelines for Development Projects in Hong Kong, If the ET Leader proposes to use a direct reading dust meter to measure 1-hr TSP levels, he shall submit sufficient information to the IC(E) to prove that the instrument is capable of achieving a comparable result as that the HVS and may be used for the 1-hr sampling. The instrument shall also be calibrated regularly, and the 1-hr sampling shall be determined periodically by HVS to check the validity and accuracy of the results measured by direct reading method.
- 1.2.2. The 1-hour TSP air quality monitoring is proposed to be performed by using portable direct reading dust meters at the air quality monitoring stations of AM1, AM2, AM3(A), AM4 and AM5 which stated in Project EM&A Manual and baseline monitoring report. The monitoring stations are listed and shown in *Table 1.1* and *Figure 1.1*.

Table 1.1 Air Monitoring Stations

		Level
Monitoring Station ID	Monitoring Location	(in terms of no. of floor)
AM1	Ah Kung Kok Fishermen Village	G/F
AM2	Block H, Kam Tai Court	Roof

W16848



Lam Environmental Services Limited

AM3(A)	Kowloon City Baptist Church Hay Nien Primary School	G/F (tentative)
AM4	Wellborn Kindergarten	G/F
AM5	The Neighbourhood Advice-Action Council Harmony Manor	Roof

- 1.2.3. The portable direct reading dust meters was calibrated at 1-year interval (For Met One AEROCET 831) or 2-years interval (For Met One BT-645) based on the requirement of manufacturer.
- 1.2.4. The portable direct reading dust meters was checked with standard equipment, High volume Sampler (HVS) at 1-year interval to ensure the accuracy and validity. The accuracy and validity checking of portable direct reading dust meters will be carried out in order to determine the conversion factor between the portable direct reading dust meters and HVS. The accuracy and validity checking is to be considered as very strong correlation and valid if the calculated correlation coefficient (r) between 0.80-1.0 (Evans, 1996).
- 1.2.5. The brand and model of the equipment to be used for this project are given in *Table 1.2*.

Equipment Brand and model **Series Number** R22584 R22586 X19295 Met One BT- 645 X19296 X19297 Portable direct reading X19298 dust meter X19299 W14016 W15448 Met One AEROCET 831 W15449

Table 1.2 Air Quality Monitoring Equipment

- 1.2.6. The proposed portable direct reading dust meters operating principal are summarized as follow (Detail refers to Appendix A):
 - Operating Principal of Met One BT-645: The Met One Instruments, Inc. model BT-645 is a type of nephelometer which measures real-time airborne TSP particulate concentration levels using the principle of forward laser light scatter. Sample air is drawn into the BT-645 and through the laser optical module, where the particulate in the sample air stream scatters the laser light through reflective and refractive properties. This scattered light is collected onto a photodiode detector at a near-forward angle, and the resulting electronic signal is processed to determine a continuous, real-time measurement of airborne particulate mass concentrations.



- Operating Particle of Met One AEROCET 831: The AEROCET 831 counts and sizes
 particles then uses proprietary algorithm to convert count data to mass
 measurements (μg/m3). Fundamentally, the AEROCET 831 calculates a volume for
 each detected particle then assigns a standard density for the conversion assigns.
- 1.2.7. The proposed portable direct reading dust meters was adopted in EM&A project such as EDO/01/2017 Environmental Team for Development of Anderson Road Quarry Site Road improvement works and were approved by IEC and EPD.
- 1.2.8. The proposed validity and accuracy checking method was adopted in EM&A project such as EDO/01/2017 Environmental Team for Development of Anderson Road Quarry Site Road improvement works and was approved by IEC and EPD.
- 1.2.9. The specification, a sample of calibration certificate and certificate of comparison check with High Volume Sampler of the proposed air quality monitoring equipment listed in Table 1 are attached in Appendix A and Appendix B.

1.3 Monitoring Methodology

1.3.1. Measuring Procedures

- (a) Check the calibration period of portable direct reading dust meter prior to monitoring (The direct reading dust meter was calibrated at 1-year interval (For Met One AEROCET 831) or 2-years interval (For Met One BT-645) and checked with High Volume Sampler at 1-year interval.
- (b) Record the site condition near / around the monitoring stations.
- (c) Install the portable direct reading dust meter to the monitoring location.
- (d) Slide the power switch to turn the power on.
- (e) Check of portable direct reading dust meter to ensure the equipment operation in normal condition.
- (f) Select the period of measurement to 60mins.
- (g) Check and set the correct time.
- (h) Select the appropriate unit display for the equipment.
- (i) Slide the power switch to turn the power off when the monitoring period ended (3 times 1 hour TSP monitoring per day).
- (j) Uninstall the portable direct reading dust meter
- (k) Collected the sampled data for analysis.

Remark: Procedures (c) to (h) may be different subject to the brands and models of portable direct reading dust

1.3.2. Maintenance and Calibration




- (a) The Portable direct reading dust meter was calibrated at 1-year interval (For Met One AEROCET 831) or 2-years interval (For Met One BT-645) based on the requirement of manufacturer.
- (b) Accuracy and validity checking of Portable direct reading dust meter will be carried out at 1-year interval in order to determine the conversion factor between the direct reading dust meter and the standard equipment, HVS.



Contract No. SPW 25/2018 Environmental Team for Relocation of Sha Tin Sewage Treatment Works to Caverns – Site Preparation and Access Tunnel Construction

Figure 1.1 Air Quality Monitoring Stations





Contract No. SPW 25/2018 Environmental Team for Relocation of Sha Tin Sewage Treatment Works to Caverns – Site Preparation and Access Tunnel Construction

Appendix A Specification of equipment

BT-645 Portable Dust Monitor

The **BT-645** Portable Dust Monitor has been designed to continuously measure and record particulate information for indoor work place and public place environments.

Key Features:

- 'Real Time' monitoring
- Simple operation
- Rugged design
- Built-in Datalogger
- Purge air system
- Low power consumption



Optical Sensor

A laser optical sensor is used to detect and measure particulate concentrations up to 100 milligrams per cubic meter. The continuous flow optical sensor is combined with purge air to ensure accurate measurements in adverse environments.

Rugged Design

The optical sensor, matched electronics and backlit display, are enclosed in a rugged metal enclosure to protect the instrument from normal daily use.

Data Storage

The internal datalogger will store over 5,000 samples. Sample history events can be viewed on the display or downloaded to a computer using Comet software.

Battery Operation

The internal battery pack provides over 10 hours of continuous operation. Battery recharge time is approximately 2.5 hours.

Optional Inlets

The standard unit is supplied with a TSP inlet. Optional sampling inlets are available for PM2.5, PM4 and PM10.

Applications:

- Public and Workplace Monitoring
- Industrial / Occupational Hygiene
- Indoor Air Quality

Note: Not recommended for continuous outdoor use.



BT-645 Portable Dust Monitor

Measurements & Specifications

OPERATING PRINCIPLE Forward light scatter laser nephelometer

PERFORMANCE

Concentration Range 0 to 100 mg/m 3 (0 to 100,000 μ g/m 3)

Sensitivity 1 µg/m³

Accuracy 5% traceable to standard with 0.6 μm PSL

Measuring Time 1, 5, 10, 15, 30 and 60 minutes

Flow Rate 2 LPM

Note: Measurement accuracy requires use of appropriate K-factor

for the material being measured.

ELECTRICAL

Light Source Laser diode, 5 mW, 670 nm

Power 14.8V Li-ion (Lithium Ion) battery pack
Battery Life Over 10 hours continuous operation

AC Adapter/Charger Li-ion battery charger, 100 – 240 VAC, 50/60Hz to 16.8 V @ 1.8 A

Communications RS-232, USB

Data Storage Over 5,000 samples

INTERFACE

Display 4 X 20 LCD

Keyboard 6-key membrane type

PHYSICAL

Size Height = 4.25'' (10.8 cm)

Width = 7.44" (18.9 cm) Thickness = 3.25" (5.4 cm)

Weight 3.7 lbs (1.7 kg)

ENVIRONMENTAL

Operating Temperature 0°C to +50°C

Relative Humidity 90% (non-condensing)

Storage Temperature -20°C to +60°C

ACCESSORIES

Supplied User Manual

USB Cable

Communications Software (Comet)
AC Adapter / Battery Charger

Optional PM2.5 Sharp Cut Cyclone (SCC-112)

PM10 Sharp Cut Cyclone (SCC-110)

Custom Serial Cable (PN 3228)

BT-645 w/SCC-112 PM2.5 Sharp Cut Cyclone



REV SEPT 2014

AEROCET 831 Aerosol Mass Monitor

Small - Reliable - Affordable

Key Features:

Five Mass Ranges

(PM1, PM2.5, PM4, PM10 and TSP)

A mass monitor that simultaneously provides five important mass ranges in one minute.

Handheld Operation

Survey the environment with this extremely portable, size selective mass monitor. This small lightweight instrument is only 28 ounces and is the perfect survey tool for a wide range of applications.

Simple Operation

Only two front-panel buttons accompany the multifunction rotary scroll wheel to provide simple and efficient operation. Just press and rotate to scroll through the menu driven platform.

Sample History

View sample history easily on the display or export data via the USB port using the included software. Unit stores up to 2500 sample events. USB interface provides quick data download and field firmware upgrades.

Battery Powered

24 hours of typical intermittent operation and up to 8 hours of continuous use. Recharge time is only 2.5 hours. The included AC charger / adapter operates the counter even when the batter pack is discharged.



Figure 1 – AEROCET 831 with optional protective boot

Applications:

- Indoor Air Quality
- Industrial / Occupational Hygiene
- HVAC Applications

Note: Not recommended for continuous outdoor use.

Avoid sampling in rainy or foggy conditions.



AEROCET 831 Aerosol Mass Monitor

Specifications:

Operating Principle Particle count to mass conversion

Features

Sample Duration 1 minute

Sample Modes Manual and Continuous

Data storage 2,500 records

Display 2-line by 16-character LCD
Controls 2 button keypad with rotary dial

Performance

PM Ranges PM1, PM2.5, PM4 and PM10

Concentration Range $0 - 1,000 \mu g/m^3$

Resolution $0.1 \mu g/m^3$ (display / serial output) Sensitivity High = $0.3 \mu m$, Low = $0.5 \mu m$ Accuracy $\pm 10\%$ to calibration aerosol

Flow rate 0.1 CFM (2.83 lpm)

Electrical

Light Source Laser Diode, 780 nm, 40 mW typical

AC Adapter/Charger AC to DC module, 100 – 240 VAC to 8.4 VDC

Battery Type Li-ion rechargeable Battery

Battery Operating Time 8 hours continuous operation. Up to 24 hours intermittent operation.

Battery Recharge Time 2.5 hours typical Communication USB Mini B Type

Physical

Weight

Size Height: 6.25" (15.9 cm) Width: 3.63" (9.22 cm)

Thickness: 2.00" (5.08 cm) 1.74 lbs – 28 ounces – (0.79 kg)

Environmental

Operating Temperature 0° C to +50° C Storage Temperature -20° C to +60° C

Accessories

Supplied Battery charger / adapter

USB Cable

Operation manual Comet software

Optional Carrying Case (PN 8517)

Protective Boot (PN 80450) Zero Filter Kit (PN 80846) Flow Meter Kit (PN 80530)

REV JUNE 2014





Contract No. SPW 25/2018 Environmental Team for Relocation of Sha Tin Sewage Treatment Works to Caverns – Site Preparation and Access Tunnel Construction

Appendix B Sample of equipment calibration certification



Certificate of Calibration

BT-645

Particulate Monitor

Recommended calibration	n interval is	24 months	from first	day of	use.
-------------------------	---------------	-----------	------------	--------	------

Accommented Co		i vai is 24 monins from f	ist may of moci
Unit Info Model:	BT-645	81865 Firmware Rev:	1.1.0
Serial Number:	R22584	81113	0.2.4
Calibrated By:	Kevin Ricks	Cal. Date:	01/18/2019
Quality Inspector	A 25	Date:	JAN 2 1 2019
Calibration Hz/µg/m³:	9.50		
Final Test			
Flow (2.0 L/M):	Pass	Ambient T (C)	22
		RH, %	34
Serial Communication:	Pass	RH, %	34

Calibration Standards

Standards	Manufacturer	Model	SN	Cal Due
RMS Multimeter	Fluke	289 Multimeter	23740018	5/03/2019
RH &TEMPERATURE	Met One Instruments	083E-1-6	R20313	9/18/2019
Primary Flow Meter	BIOS	Defender-510	1033419	3/28/2019
Digital Dust Indicator	SIBATA	LD-3B	476795	5/18/2019

The standards used for this calibration have accuracy equal to or greater than the instrument tested. These standards are on record and traceable to NIST to the extent allowed by the institute's calibration facility. Unless otherwise stated, all instruments are calibrated to meet the manufacturer's published specifications. The Calibration system complies with MIL-STD-45662A.



Portable Dust Meter Performance Check Record

Portable Dust Meter

Type : Particulare Monitor

Manufacturer : MET ONE INSTRUMENTS

Model Number : BT-645

Serial Number : R22584

Performance Check Date : 27-Feb-19

Standard Equipment

Type : High Volume Sampler

Manufacturer : TISCH

Model Number : TE-5170

Equipment Number : HVS018


Last Calibration Date : 4-Dec-18

Portable Dust Meter Performance Check Results

Trial no. in 1-hr period	Time	Mean Pressure (hPa)	Mean Temp (°C)	Concentration in ug/m ³ (Standard equipment) (Y - Axis)	Concentration in ug/m ³ (Performance Check / Calibrated equipment) (X - Axis)
Zero Check	27/2/19 07:00	1016	21	0	0
1	27/2/19 08:45	1016	21	75	61
2	27/2/19 09:52	1016	21	53	51
3	27/2/19 11:00	1016	21	72	58

^{*} Filter paper weighting was conducted by HOKLAS accredited laboratory.

Linear Regression of Y on X



Checked by: Chan Ka Chun Date: 4-Mar-19	Operator:	Henry Lau	Date:	27-Feb-19
	Checked by:	Chan Ka Chun	Date:	4-Mar-19



Calibration Certificate

The calibration results on this report certify that this instrument complies with the product specifications at the time of calibration. Calibration was performed according to accepted industry methods using equipment, procedures, and standards that are traceable to NIST and ISO.

Instrument Model# Aerocet 831	1	Instrument Serial#	W15449
Date of Calibration 10/4/2018			Sensor # 16439
Darleen Best		A 25	
Calibration Technician		Quality Check	
Temperature 23	°C	Relative Humidity 36	6.5 %

Recommended calibration interval is 12 months from the first day of use.

Test Procedure: Aerocet 831-6100

Test Results	Test Spec.	Lot# NIST	Expiration
Pass	± 10%	183039	03/31/2020
Pass	± 10%	180556	02/28/2020
Pass	± 10%	169240	5/31/2019
Pass	± 10%	REF	, NA
Pass	± 10%	REF	NA
Pass	± 10%	REF	NA
Pass	± 10%	REF	NA
Pass	± 10%	REF	NA
	Pass Pass Pass Pass Pass Pass Pass	Pass ± 10% Pass ± 10%	Pass ± 10% 183039 Pass ± 10% 180556 Pass ± 10% 169240 Pass ± 10% REF Pass ± 10% REF Pass ± 10% REF Pass ± 10% REF Pass ± 10% REF

Standards	Model	SN	Cal Due
Particle Counter	GT-526	M1760	10/9/2018
Flowmeter	DCL-M	103751	1/29/2019
DMM	289	27720071	6/29/2019
RH/TEMP SENSOR	083E-1-6	R20313	9/18/2019

This calibration certificate shall not be reproduced except in full, without the written approval of Met One Instruments Inc.



REPORT OF EQUIPMENT PERFORMANCE CHECK / CALIBRATION

REPORT NO. PROJECT NAME DATE OF ISSUE

HK1811054 PERFORMANCE CHECK / CALIBRATION OF DUST METER

24/10/2018

: LAM ENVIRONMENTAL SERVICES LTD CUSTOMER

: 11/F, CENTRE POINT, 181-185 GLOUCESTER ROAD, WAN CHAI, HONG KONG **ADDRESS**

REPORT NO. HK1811054 PROJECT ITEM NO. HK1811054-01

PERFORMANCE CHECK / CALIBRATED EQUIPMENT

AEROSOL MASS MONITOR MANUFACTURER MET ONE INSTRUMENTS

MODEL NO. SERIAL NO. AEROCET - 831 W15449 **EQUIPMENT NO.** 18/10/2018 RECEIPT DATE

PERFORMANCE CHECK / CALIBRATION DATE : 23/10/2018 PERFORMANCE CHECK / CALIBRATION Information

CODE	Calibration Parameter	Method Procedure	Reference Method
Dust PC/CAL	Performance Check / Calibration of Dust Meter	CAL003	General Technical Requirements of Environmental Monitoring, Environmental Monitoring & Audit Guidelines for Development Projects in HK

Notes: 1. This report shall not be reproduced, except in full, without prior approval from Pilot Testing Limited.

2. Performance Check / Calibration result relates to performance check / calibration item(s) as received.

Approved Signatory

Wong Po Yan Pauline (Assistant Laboratory Manager) Issue Date:

24/10/2018



REPORT OF PERFORMANCE CHECK / CALIBRATION

PERFORMANCE CHECK / CALIBRATION OF DUST METER 24/10/2018 PROJECT NAME

DATE OF ISSUE REPORT NO. HK1811054

PERFORMANCE CHECK / CALIBRATED EQUIPMENT

AEROSOL MASS MONITOR **MANUFACTURER** MET ONE INSTRUMENTS

MODEL NO. AEROCET - 831

SERIAL NO. W15449 EQUIPMENT NO.

PERFORMANCE CHECK / CALIBRATION DATE 23/10/2018

STANDARD EQUIPMENT

TYPF HIGH VOLUME AIR SAMPLER

MANUFACTURER TISCH MODEL NO. TE-5170 EQUIPMENT REF NO. PTL_HV002 LAST CALIBRATION DATE 25/7/2018

EQUIPMENT PERFORMANCE CHECK / CALIBRATION RESULTS:


Trial no. in 1-hr period	Time	Mean Temp (°C)	Mean Pressure (hPa)	Concentration in ug/m³ (Standard equipment) (Y - Axis)	Concentration in ug/m³ (Performance Check / Calibrated equipment) (X - Axis)
Zero Check ¹	23/10/2018,9:05:00 AM	25.3	1017	0	0
1	23/10/2018,10:20:00 AM	25.3	1017	45	51
2	23/10/2018,11:22:00 AM	25.3	1017	32	39
3	23/10/2018,12:29:00 PM	25.3	1017	12	19

Linear Regression of Y on X

Slope (K- factor) Correlation Coefficient

Validity of Performance Check / Calibration Record

23/10/2019



Notes: 1. Zero check conducted as per CAL003 SOP and manufacturer's manual as appropriate.

2. This report shall not be reproduced, except in full, without prior approval from Pilot Testing Limited.

3. Performance Check / Calibration result relates to performance check / calibration item(s) as received.

Operator:	Lau, Natalie	Signature:	letter	Date:	23/10/2018

Checked by: Wong Po Yan, Pauline Signature: 24/10/2018 Date:



Contract No. SPW 25/2018 Environmental Team for Relocation of Sha Tin Sewage Treatment Works to Caverns – Site Preparation and Access Tunnel Construction

Appendix C Project layout plan

