

**High-Volume TSP Sampler**

**1-Point Calibration Record**

Location : A1  
Calibrated by : K.F.Ho  
Date : 20/6/08

**Sampler**

Model : GMWS-2310 ACCU-VOL  
Serial Number : S/N 9315

**Calibration Office and Standard Calibration Relationship**

Serial Number : CM-AIR-43  
Service Date : 2 July 2007  
Slope (m) : 0.057452  
Intercept (b) : -0.026137  
Correlation Coefficient(r) : 0.999910

**Standard Condition**

Pstd (hpa) : 1013  
Tstd (K) : 298.18

**Calibration Condition**

Pa (hpa) : 1012  
Ta(K) : 301

**IC (Indicated flow)** : 38 cfm

**Actual flow** : 1.21 m<sup>3</sup>/min

Checked by: Magnum Fan

Date: 23/06/08

High-Volume TSP Sampler  
5-Point Calibration Record

Location : A1  
 Calibrated by : K.T.Ho (Kennedy Town)  
 Date : 21/04/08

Sampler  
 Model : GMWS-2310 ACCU-VOL  
 Serial Number : S/N 9315

Calibration Office and Standard Calibration Relationship

Serial Number : CM-AIR-43  
 Service Date : 2 July 2007  
 Slope (m) : 0.057452  
 Intercept (b) : -0.026137  
 Correlation Coefficient(r) : 0.999910

Standard Condition

Pstd (hpa) : 1013  
 Tstd (K) : 298.18

Calibration Condition

Pa (hpa) : 1011  
 Ta(K) : 298

Resistance Plate	d11 [green liquid] (inch water)	Z	X=Qstd (cubic meter/min)	IC	Y
1	18 holes	3.469	1.688	58	58.3
2	13 holes	3.148	1.533	50	50.3
3	10 holes	2.735	1.333	42	42.2
4	7 holes	2.179	1.065	30	30.2
5	5 holes	1.683	0.825	18	18.1

Sampler Calibration Relationship

Slope(m): 45.839 Intercept(b): -19.249 Correlation Coefficient(r): 0.9994

Checked by: Magnum Fan

Date: 21/04/08

**High-Volume TSP Sampler**

**1-Point Calibration Record**

Location : A2a  
Calibrated by : K.F.Ho  
Date : 20/06/08

**Sampler**

Model : GMWS-2310 ACCU-VOL  
Serial Number : S/N 9048

**Calibration Office and Standard Calibration Relationship**

Serial Number : CM-AIR-43  
Service Date : 2 July 2007  
Slope (m) : 0.057452  
Intercept (b) : -0.026137  
Correlation Coefficient(r) : 0.999910

**Standard Condition**

Pstd (hpa) : 1013  
Tstd (K) : 298.18

**Calibration Condition**

Pa (hpa) : 1012  
Ta(K) : 301

**IC (Indicated flow)** : 38 cfm

**Actual flow** : 1.25 m<sup>3</sup>/min

Checked by: Magnum Fan

Date: 23/06/08

High-Volume TSP Sampler  
5-Point Calibration Record

Location : A2(Kennedy Town)  
Calibrated by : K.T.Ho  
Date : 21/04/08

Sampler

Model : GMWS-2310 ACCU-VOL  
Serial Number : S/N 9048

Calibration Office and Standard Calibration Relationship

Serial Number : CM-AIR-43  
Service Date : 2 July 2007  
Slope (m) : 0.057452  
Intercept (b) : -0.026137  
Correlation Coefficient(r) : 0.999910

Standard Condition

Pstd (hpa) : 1013  
Tstd (K) : 298.18

Calibration Condition

Pa (hpa) : 1011  
Ta(K) : 298

Resistance Plate	dH [green liquid] (inch water)	Z	X=Qstd (cubic meter/min)	IC (indicated flow)	Y
1   18 holes	10.5	3.232	1.573	56	55.9
2   13 holes	8.0	2.821	1.375	48	47.9
3   10 holes	6.3	2.504	1.221	40	39.9
4   7 holes	3.9	1.970	0.963	26	25.9
5   5 holes	2.4	1.545	0.758	12	12.0

Sampler Calibration Relationship


Slope(m): 53.934 Intercept(b): -27.24

Correlation Coefficient(r): 0.9959

Checked by: Magnum Fan

Date: 21/04/08





輝創工程有限公司

Sun Creation Engineering Limited Calibration and Testing Laboratory

Certificate No. : C081909

## Certificate of Calibration

*This is to certify that the equipment*

*Description : Sound Level Meter*

*Manufacturer : Rion*

*Model No. : NL-31*

*Serial No. : 00410224*

*has been calibrated for the specific items and ranges.  
The results are shown in the Calibration Report No. C081909.*

*The equipment is supplied by*

*Co. Name : Envirotech Services Co.*

*Address : Shop 6, G/F., Casio Mansion, 209 Shaukeiwan Road,  
Hong Kong*

*Date of Issue : 14 April 2008*


Certified by :

*K C Lee*

The test equipment used for testing are traceable to the National Standards as specified in this report.  
This report shall not be reproduced except in full and with prior written approval from this laboratory.

Calibration and Testing Laboratory of Sun Creation Engineering Limited

c/o 4/F. Tsing Shan Wan Exchange Building, 1 Hing On Lane, Tuen Mun, New Territories, Hong Kong  
Tel: 2927 2606 Fax: 2714 8986 E-mail: callab@suncreation.com Website: www.suncreation.com



輝創工程有限公司

Sun Creation Engineering Limited Calibration and Testing Laboratory

Certificate No. : C083194

## Certificate of Calibration

*This is to certify that the equipment*

*Description : Sound Level Calibrator*

*Manufacturer : Rion*

*Model No. : NC-73*

*Serial No. : 10997142*

*has been calibrated for the specific items and ranges.  
The results are shown in the Calibration Report No. C083194.*

*The equipment is supplied by*

*Co. Name : Envirotech Services Co.*

*Address : Shop 6, G/F., Casio Mansion, 209 Shauketwan Road,  
Hong Kong*

*Date of Issue : 25 June 2008*

*Certified by:*

*K. C. Lee*

The test equipment used for testing are traceable to the National Standards as specified in this report.  
This report shall not be reproduced except in full and with prior written approval from this laboratory.

Calibration and Testing Laboratory of Sun Creation Engineering Limited

c/o 4/F. Tsing Shan Wai Exchange Building, 1 Hing On Lane, Tuen Mun, New Territories, Hong Kong

Tel. 2927 2606

Fax. 2744 8986

E-mail: calllab@suncreation.com

Website: www.suncreation.com