

## **Appendix I**

### **Calibration Certificate of Portable H<sub>2</sub>S Meter**



**mPower Electronics**

*Making Powerful Senses*

### Calibration and Test Certificate

**Product Name:** POLI  
**Model Number:** MP400P  
CO2 0-5000ppm  
LEL 0-100%LEL  
O2 0-30%  
H2S 0-100ppm  
**Serial Number:** M00401003200  
**Inspection Date:** 1/27/2022

#### Calibration and Test Gases

#	Gas	Concentration	Lot#
1	CO2	5000ppm	L74402180
2	CO	60ppm	93401123
	H2S	15ppm	
	O2	18%	
	LEL	50%LEL(2.5%VOL)	
3	N2	99.9%	2005012

#### Test Results:

#	Sensor	Span	UOM
1	CO2	5020	ppm
2	LEL	49	%LEL
3	O2(18% / 0%)	18.0 / 0.0	% / %
4	H2S	15.1	ppm

*This instrument has been calibrated using valid calibration gases and instrument manual operation procedures. Test and calibration data is on file with the manufacturer, mPower Electronics.*

Approved By: *Huiying Yao*

**mPower Electronics, Inc.**  
3046 Scott Boulevard, Santa Clara, CA 95054  
Phone: (408)320-1266 Fax: (669)342-7077  
www.mpowerinc.com